

# SensiMix™ SYBR Hi-ROX One-Step Kit

Shipping: On Dry/Blue Ice	Catalog numbers
Exp. Date: See vial	QT245-01: 25 x 50µl reactions: 1 x 625µl
Batch No.: See vial	QT245-02: 250 x 50µl reactions: 5 x 1.25ml
Concentration: see vial	QT245-05: 500 x 50µl reactions: 10 x 1.25ml
	QT245-20: 2000 x 50µl reactions: 40 x 1.25ml



A Meridian Life Science® Company

Store at -20°C

## Storage and stability:

The SensiMix™ SYBR Hi-ROX One-Step Kit is shipped on Dry/Blue Ice. All kit components should be stored at -20°C upon receipt. Excessive freeze/thawing is not recommended. When stored under optimum conditions, the reagents are stable for a minimum of 12 months from date of purchase.

## Quality control:

Bioline operates under ISO 9001 Management System. The SensiMix™ SYBR Hi-ROX One-Step Kit and its components are extensively tested for activity, processivity, efficiency, heat activation, sensitivity, absence of nuclease contamination and absence of nucleic acid contamination.

## Safety precautions:

Harmful if swallowed. Irritating to eyes, respiratory system and skin. Please refer to the material safety data sheet for further information.

## Description

The SensiMix™ SYBR Hi-ROX One-Step Kit contains a ready-to-use reagent for real-time RT-PCR for use on various real-time instruments. The SensiMix™ SYBR Hi-ROX One-Step Kit provides high-performance, reproducible first strand cDNA synthesis and subsequent SYBR Green-based real-time PCR assays, using a single protocol. Both cDNA synthesis and PCR are performed in a single tube using gene-specific primers and either total RNA or mRNA. For high specificity and sensitivity of the one-step RT-PCR, SensiMix™ SYBR Hi-ROX One-Step Kit comes with a separate tube of reverse transcriptase which is added into a mix containing a proprietary hot-start DNA polymerase. SensiMix One-Step Kit is inactivated and possesses no polymerase activity during the reaction setup, preventing non-specific amplification including primer-dimer formation.

The SensiMix™ SYBR Hi-ROX One-Step Kit consists of a 2x SensiMix™ SYBR Hi-ROX One-Step mix, separate reverse transcriptase and RiboSafe RNase Inhibitor.

## Kit components

Reagent	25 x 50µl Reactions	250 x 50µl Reactions	500 x 50µl Reactions	2000 x 50µl Reactions
SensiMix™ SYBR Hi-ROX One-Step (2x)	1 x 625µl	5 x 1.25ml	10 x 1.25ml	40 x 1.25ml
50mM MgCl <sub>2</sub>	1 x 1ml	1 x 1ml	1 x 1ml	4 x 1ml
RNase Inhibitor (10U/µl)	1 x 25µl	1 x 250µl	1 x 500µl	4 x 500µl
100x Reverse Transcriptase	1 x 12.5µl	1 x 125µl	1 x 250µl	1 x 1ml
DEPC-H <sub>2</sub> O	1 x 1.8ml	2 x 1.8ml	4 x 1.8ml	16 x 1.8ml

## Kit compatibility

The SensiMix™ SYBR Hi-ROX One-Step Kit has been optimized for use in SYBR Green-based real-time RT-PCR on the real-time instruments listed in the following compatibility table, each of these instruments having the capacity to analyze the real-time PCR data with the passive reference signal either on or off. The kit is also compatible with several instruments that do not require the use of ROX, such as the Qiagen (Corbett) Rotor-Gene™ 6000, the Bio-Rad CFX96 or the Roche LightCycler® 480.

Manufacturer	Model
ABI	7000, 7300, 7700, 7900, 7900HT, StepOne™ and StepOne™ Plus

## General considerations

When handling RNA it is important to use RNase-free plasticware and reagents. We also recommend performing RNA work in an isolated area. To help prevent any carry-over DNA contamination, we recommend that separate areas are maintained for PCR set-up, PCR amplification and any post-PCR gel analysis. It is essential that any tubes containing amplified PCR product are not opened in the PCR set-up area.

**Primers:** The sequence and concentration of the primers, as well as amplicon length, can be critical for specific amplification, yield and overall efficiency of any real-time PCR. We strongly recommend taking the following into consideration when designing and running your PCR reaction:

- use primer-design software, such as Primer3 or visual OMP™ (<http://frodo.wi.mit.edu/primer3/> and DNA Software, Inc. <http://dnasoftware.com/>, respectively). Primers should have a melting temperature (T<sub>m</sub>) of approximately 60°C.
- optimal amplicon length should be 50-150bp, and should not exceed 400bp
- final primer concentration of 250nM is suitable for most PCR conditions, however to determine the optimal concentration we recommend titrating in the range of 0.1-1µM
- use an equimolar primer concentration

**Template:** It is important that the RNA template is intact and devoid of contaminating inhibitors for both reverse transcription and subsequent PCR. For high purity RNA, we recommend using the Bioline ISOLATE RNA Mini Kit (BIO-52043). Prior to use in RT-PCR, RNA suspensions and dilutions should be made in DEPC-treated Water (BIO-38030), to avoid any RNase-mediated degradation.

The recommended amount of template for one-step RT-PCR is dependent upon the type of RNA used. The following should be considered when using total RNA and isolated mRNA:

- **total RNA:** purified total RNA can be used in the range from 1pg to 1µg per 50µl reaction
- **mRNA:** purified mRNA can be used from 0.5pg per 50µl reaction

**MgCl<sub>2</sub>:** The MgCl<sub>2</sub> concentration in the 1x reaction mix is 3mM, which is optimal for both reverse transcriptase and SensiAq in the majority of real-time RT-PCR conditions. If necessary, we recommend titrating MgCl<sub>2</sub> to a maximum of 5mM.

**RT-PCR controls:** It is important to detect the presence of contaminating genomic DNA that may affect the reliability of the data. Always include a no-template control, replacing the template with PCR-grade water and a no-Reverse Transcriptase control to ensure the RNA sample is not contaminated with DNA.

**Optional ROX :** The SensiMix SYBR Hi-ROX One-Step Kit is premixed with ROX (5-carboxy-X-rhodamine, succinimidyl ester), so that where necessary, ROX fluorescence can be optionally detected on certain real-time instruments. If your real-time instrument has the capability of using a ROX and you wish to use this option, then this option must be selected by the user in the software (see notice to purchaser No. 5 in Trademark and Licensing Information).

## Procedure

**Reaction mix composition:** Prepare a master mix. The volumes given below are based on a standard 50µl final reaction mix and can be scaled accordingly.

Reagent	Volume	Final concentration
2x SensiMix SYBR Hi-ROX One-Step	25µl	1x
25µM Forward Primer	0.5µl	250nM
25µM Reverse Primer	0.5µl	250nM
100x Reverse Transcriptase	0.5µl	1x
10U/µl RiboSafe RNase Inhibitor	1µl	0.2U/µl
DEPC-H <sub>2</sub> O	Up to 45µl	-
Template	5µl	
<b>50µl Final volume</b>		

## Troubleshooting guide

Problem	Possible Cause	Recommendation
No amplification trace AND No product on agarose gel	Activation time too short	Make sure SensiMix is activated for 10min at 95°C before cycling
	Error in protocol setup	Verify that correct reagent concentrations, volumes, dilutions and storage conditions have been used
	Suboptimal primer design	Use primer design software or validated primers. Test primers on a control template
	Incorrect concentration of primers	Use primer concentration between 100nM and 1µM
	Template degraded	Re-isolate your template from the sample material or use freshly prepared template dilution
	Primers degraded	Use newly synthesized primers
	Template contaminated with PCR inhibitors	Further dilute template before PCR or purify template and resuspend it in PCR-grade H <sub>2</sub> O
	Template concentration too low	Increase concentration used
	Cycling conditions not optimal	Increase extension/annealing times, increase cycle number, reduce annealing temperature
No amplification trace AND Product on agarose gel	Error in instrument setup	Check that the acquisition settings are correct during cycling

**Suggested RT-PCR conditions:** The following RT-PCR conditions are suitable for SensiMix SYBR Hi-ROX One-Step Kit with a majority of amplicons and real-time PCR instruments. However, the cycling conditions can be varied to suit different machine-specific protocols. The critical step of the RT-PCR is the 10 minute polymerase activation at 95°C. The detection channel on the real-time instrument should be set to SYBR Green or FAM.

The following RT-PCR conditions are suitable for SensiMix SYBR Hi-ROX One-Step Kit with a majority of amplicons and real-time PCR instruments.

Cycles	Temperature	Time	Notes
1	45°C	10min	Reverse transcription
1	*95°C	*10min	Polymerase activation
40	95°C	15sec	Temperature dependent on T <sub>m</sub> of primers Acquire at end of step
	60°C	15sec	
	72°C	15sec	

\*Non-variable parameter

**Optional analysis:** After the reaction has reached completion, refer to the instrument instructions for the option of melt-profile analysis.

## Troubleshooting guide (continued)

Problem	Possible Cause	Recommendation
Non-specific amplification product AND Primer-dimers	Suboptimal primer design	Redesign primers using appropriate software or use validated primers
	Primer concentration too high	Test dilution series of primer concentrations until primer dimer/non-specific amplification products disappear
	Primer concentration too low	Titrate primers in the concentration range of 100nM - 1 $\mu$ M
	Primer annealing temperature too low	Increase PCR annealing temperature in increments of 2°C until primer dimer/non-specific amplification products disappear
	Template concentration too low	Increase template concentration
	Template concentration too high	Reduce template concentration until non-specific products disappear
	Extension time too long	Reduce extension time to determine whether non-specific products are reduced
Late amplification trace	Activation time too short	Ensure that the reaction is activated for 10min at 95°C before cycling
	Annealing temperature too high	Decrease annealing temperature in steps of 2°C
	Extension time too short	Double extension time to determine whether the cycle threshold (C <sub>T</sub> ) is affected
	Template concentration too low	Increase concentration if possible
	Template is degraded	Re-isolate template from sample material or use freshly prepared template dilution
	Suboptimal design of primers	Redesign primers using appropriate software or use validated primers
	Primer concentration too low	Increase concentration of primer in 100nM increments
PCR efficiency below 90%	Extension time is too short	Increase extension time
	Primer concentration too low	Increase concentration of primer in 100nM increments
	Suboptimal design of primers	Redesign primers using appropriate software or use validated primers
PCR efficiency above 110%	Template is degraded or contains PCR inhibitors	Re-isolate template from sample material or use freshly prepared template dilution or purify template and resuspend it in H <sub>2</sub> O
	Non specific amplification and/or primer-dimers	Use melt analysis and 4% agarose gel electrophoresis to confirm presence of non-specific amplification products. See above for preventing/removing non-specific products

## Technical Support

If the troubleshooting guide does not solve the difficulty you are experiencing, please contact Technical Support with details of reaction setup, cycling conditions and relevant data.

Email: [tech@bioline.com](mailto:tech@bioline.com)

## Associated Products

Product	Description	Pack Size	Cat No.
ISOLATE Genomic DNA Mini kit	Rapid isolation of DNA from a variety of samples	10 Preps 50 Preps 250 Preps	BIO-52031 BIO-52032 BIO-52033
ISOLATE Plant DNA Mini kit	Rapid isolation of DNA from a variety of plant samples	10 Preps 50 Preps 250 Preps	BIO-52034 BIO-52035 BIO-52036
ISOLATE RNA Mini Kit	Fast and efficient isolation of extremely pure total RNA from a variety of samples	10 Preps 50 Preps 250 Preps	BIO-52039 BIO-52040 BIO-52041
ISOLATE Plant RNA Mini Kit	Fast and efficient isolation of extremely pure total RNA from a variety of plant samples	10 Preps 50 Preps 250 Preps	BIO-52042 BIO-52043 BIO-52044
TRIsure™	Quick isolation of high-quality RNA from a variety of sources for subsequent use in cDNA synthesis	100ml 200ml	BIO-38032 BIO-38033
Tetro cDNA Synthesis Kit	Fully optimized to generate maximum yields of full-length cDNA from RNA	30 Reactions 100 Reactions	BIO-65042 BIO-65043
Agarose	Molecular biology grade agarose	100g 500g	BIO-41026 BIO-41025
PCR Water	Ultra-pure (18.2MΩ) molecular biology grade water	10 x 10ml	BIO-37080
DEPC-treated Water	Deionized, high-quality molecular grade water treated with DEPC. Ideal for use in all RNA work	10 x 10ml 1 Liter	BIO-38030 BIO-38031

### TRADEMARK AND LICENSING INFORMATION

- 1) Trademarks: SensiMix™ (Bioline Reagents Ltd), SYBR® (Molecular Probes), ROX™, StepOne™ (ABI), Rotor-Gene™ (Qiagen), LightCycler® (Roche).
- 2) Purchase of this product includes limited right to use the supplied amount of SYBR® Green I Stain patented by Molecular Probes, Inc.
- 3) Notice to Purchaser: Limited License. Use of this product may be covered by one or more of the following US patents: 6,127,155, 5,677,152 (claims 1 to 23 only), 5,773,258 (claims 1 and 6 only). The purchase of this product includes a limited, non-transferable immunity from suit under the foregoing patent claims for using only this amount of product for the purchaser's own internal research. No right to perform commercial services of any kind, including without limitation reporting the results of purchaser's activities for a fee or other commercial consideration, is conveyed expressly, by implication, or by estoppel. This product is for research use only. Diagnostic uses under Roche patents require a separate license from Roche. Further information on purchasing licenses may be obtained by contacting the Director of Licensing, Applied Biosystems, 850 Lincoln Centre Drive, Foster City, California 94404, USA.
- 4) SensiMix™ products are manufactured by Bioline Reagents Ltd.
- 5) Notice to Purchaser: No rights are conveyed with respect to US patent 5,928,907

Bioline Ltd  
UNITED KINGDOM  
  
Tel: +44(0)20 8830 5300  
Fax: +44 (0)20 8452 2822

Bioline USA Inc.  
USA  
  
Tel: +1 508 880 8990  
Fax: +1 508 880 8993

Bioline GmbH  
GERMANY  
  
Tel: +49(0)33 7168 1229  
Fax: +49 (0)337168 1244

Bioline (Aust) Pty. Ltd  
AUSTRALIA  
  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763