


| | | |
|--------------------------------------------------------------------------------------------------------------------------------------|----------------------------------|--------------------|
|  <small>A Mediatech Life Sciences Company</small> | <h2>Certificate of Analysis</h2> | COA No: CA BB-0054 |
| | | Version: 02 |

| | | |
|----------------------------------------------------------------------------|---------------------|--------------|
| <h1>Solution A MyTaq™ Extract Kit</h1> <p>For Research Use Only</p> | Storage Conditions: | -20°C |
| | Lot number: | MXA-616112 |
| | Expiry date: | January 2019 |

Quality Control Parameters

| Analysis | Specification | Result |
|---------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------|
| Functional | Fragment of size 1Kb was amplified, with a dilution series of DNA extracted from mouse tails and mouse cells, using standard conditions and 35 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained). | Passed |
| DNA contamination | Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample. | Passed |
| DNase contamination | Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase. | Passed |

Authorised by Christopher Weatherall



United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

Germany

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

France

Info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Singapore

Info.sg@bioline.com
Tel: +65 6774 7196
Fax: +65 6774 6441



Certificate of Analysis

COA No: CA BB-0055

Version: 02

Solution B MyTaq™ Extract Kit

For Research Use Only

Storage Conditions: -20°C

Lot number: MXB-616112

Expiry date: January 2019

Quality Control Parameters

| Analysis | Specification | Result |
|---------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------|
| Functional | Fragment of size 1Kb was amplified, with a dilution series of DNA extracted from mouse tails and mouse cells, using standard conditions and 35 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained). | Passed |
| DNA contamination | Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample. | Passed |
| DNase contamination | Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase. | Passed |

Authorised by Christopher Weatherall

United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

Germany

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

France

Info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Singapore

Info.sg@bioline.com
Tel: +65 6774 7196
Fax: +65 6774 6441



Certificate of Analysis

COA No: CA BMM-0020

Version: 02

MyTaq™ HS Red Mix

For Research Use Only

Storage Conditions: -20°C

Lot number: MTHRX-616212

Expiry date: January 2019

Quality Control Parameters

| Analysis | Specification | Result |
|---------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------|
| Functional | Fragments of sizes 525bp, 750bp, 900bp and 1300bp are amplified with a dilution series of human genomic DNA, using standard conditions and 35 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained). | Passed |
| DNA contamination | Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample. | Passed |
| DNase contamination | Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase. | Passed |

Authorised by Christopher Weatherall

United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

Germany

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

France

Info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Singapore

Info.sg@bioline.com
Tel: +65 6774 7196
Fax: +65 6774 6441