

Associated products

Product	Description	Pack Size	Cat No.
ISOLATE II Genomic DNA Kit	Rapid isolation of high-quality genomic DNA from a wide variety of samples	10 Preps 50 Preps 250 Preps	BIO-52065 BIO-52066 BIO-52067
ISOLATE II Plant DNA Kit	Rapid isolation of high-quality genomic DNA from a wide variety of plant species	10 Preps 50 Preps 250 Preps	BIO-52068 BIO-52069 BIO-52070
ISOLATE II RNA Mini Kit	Isolation of high-yield and extremely pure total RNA from a variety of samples	10 Preps 50 Preps 250 Preps	BIO-52071 BIO-52072 BIO-52073
ISOLATE II RNA Plant Kit	Isolation of high-yield and extremely pure total RNA from a wide variety of plant species	10 Preps 50 Preps	BIO-52076 BIO-52077
TRIsure™	Quick isolation of high-quality RNA from a variety of sources for subsequent use in cDNA synthesis	100 mL 200 mL	BIO-38032 BIO-38033
SensiFAST cDNA Synthesis Kit	Fully optimized to generate maximum yields of full-length cDNA from RNA	50 Reactions 250 Reactions	BIO-65053 BIO-65054
Agarose	Molecular biology grade agarose	100 g 500 g	BIO-41026 BIO-41025

Technical support

If the troubleshooting guide does not solve the difficulty you are experiencing, please contact Technical Support with details of reaction setup, cycling conditions and relevant data.

Email: mbi.tech@meridianlifescience.com

Trademark and licensing information

- 1) Trademarks: SensiFAST™ (Bioline Reagents Ltd), SYBR® (Molecular Probes), StepOne™ (ABI), RotorGene™ (Qiagen), LightCycler®, ROX™ (Roche).
- 2) Purchase of this product conveys a licence from Life Technologies to use this SYBR® containing reagent in an end-user RUO assay. Parties wishing to incorporate this SYBR® containing reagent into a downstream kit, should contact Life Technologies for SYBR® Licensing information.

Bioline Reagents Ltd
UNITED KINGDOM
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

Bioline USA Inc.
USA
Tel: +1 901.382.8716
Fax: +1 901.382.0027

Bioline GmbH
GERMANY
Tel: +49 (0) 33771 60222 00
Fax: +49 (0)3371 60222 01

Bioline (Aust) Pty. Ltd
AUSTRALIA
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Storage and Stability:
The SensiFAST SYBR® Hi-ROX Kit is shipped on dry/blue ice. All kit components should be stored at -20°C upon receipt. Excessive freeze/thawing is not recommended.

Expiry:
When stored under the recommended conditions and handled correctly, full activity of the kit is retained until the expiry date on the outer box label.

Quality control:
The SensiFAST SYBR® Hi-ROX Kit and its components are extensively tested for activity, processivity, efficiency, heat activation, sensitivity, absence of nuclease contamination and absence of nucleic acid contamination.

Safety precautions:
Please refer to the material safety data sheet for further information.

Notes:
This reagent has been manufactured under 13485 Quality Management System and is suitable for research and/or further manufacturing use.

Description

The SensiFAST™ SYBR® Hi-ROX Kit has been developed for fast, highly reproducible real-time PCR and has been validated on commonly used real-time PCR instruments. A combination of the latest advances in buffer chemistry and enhancers, together with an antibody-mediated hot-start DNA polymerase system, ensures that the SensiFAST SYBR® Hi-ROX Kit delivers fast, highly-specific and ultra-sensitive real-time PCR.

For ease-of-use and added convenience, SensiFAST SYBR® Hi-ROX is provided as a 2x mastermix containing all the components necessary for real-time PCR, including the SYBR® Green I dye, dNTPs, stabilisers and enhancers. The kit consists of a ready-to-use premix, only primers and template need to be added.

Kit components

Reagent	500 x 20 µL reactions	2000 x 20 µL reactions
SensiFAST SYBR® Hi-ROX mix (2x)	5 x 1 mL	4 x 5 mL

Instrument compatibility

SensiFAST SYBR® Hi-ROX Kit has been optimized for use in SYBR® Green-based real-time PCR on the real-time PCR instruments listed in the following compatibility table, each of these instruments having the capacity to analyze the real-time PCR data with the passive reference signal either on or off. The kit is also compatible with several instruments that do not require the use of ROX, such as the BMS Mic, Qiagen (Corbett) RotorGene™ 6000, Bio-Rad CFX96 or Roche LightCycler® 480.

Manufacturer	Model
ABI (Invitrogen)	7000, 7300, 7700, 7900, 7900HT, 7900HT FAST, StepOne™, StepOne™ Plus

General considerations

To help prevent any carry-over DNA contamination, we recommend that separate areas are maintained for reaction set-up, PCR amplification and any post-PCR gel analysis. It is essential that any tubes containing amplified PCR product are not opened in the PCR set-up area.

Primers: The specific amplification, yield and overall efficiency of any real-time PCR can be critically affected by the sequence and concentration of the primers, as well as by the amplicon length. We strongly recommend taking the following points into consideration when designing and running your real-time PCR:

- use primer-design software, such as Primer3 (<http://frodo.wi.mit.edu/primer3/>) or visual OMP™ (<http://dnasoftware.com/>). Primers should have a melting temperature (T_m) of approximately 60 °C
- optimal amplicon length should be 80-200 bp, and should not exceed 400 bp

SensiFAST™ SYBR® Hi-ROX Kit

Shipping: On dry/blue ice Catalog numbers
Batch No.: See vial BIO-92005: 500 x 20 µL reactions: 5 x 1 mL
Concentration: See vial BIO-92020: 2000 x 20 µL reactions: 4 x 5 mL

bioline
meridian BIOSCIENCE

Store at -20°C

- final primer concentration of 400 nM is suitable for most SYBR® -Green based reactions, however to determine the optimal concentration we recommend titrating in the range 0.1-1 µM. The forward and reverse primers concentration should be equimolar

- when amplifying from cDNA, use of intron spanning primers to is preferable, to avoid amplification from genomic DNA

Template: it is important that the DNA template is suitable for use in PCR in terms of purity and concentration. In addition, the template needs to be devoid of any contaminating PCR inhibitors (e.g. EDTA). The recommended amount of template for PCR is dependent upon the type of DNA used. The following points should be considered when using genomic DNA and cDNA templates:

- **Genomic DNA:** use up to 1 µg of complex (e.g. eukaryotic) genomic DNA in a single PCR. We recommend using the Bioline ISOLATE II Genomic DNA Kit (BIO-52066) for high yield and purity from both prokaryotic and eukaryotic sources.

- **cDNA:** the optimal amount of cDNA to use in a single PCR is dependent upon the copy number of the target gene. We suggest using 100 ng cDNA per reaction, however it may be necessary to vary this amount. To perform a two-step RT-PCR, we recommend using the SensiFAST cDNA Synthesis Kit (BIO-65053) for reverse transcription of the purified RNA. For high yield and purity of RNA, use the Bioline ISOLATE II RNA Mini Kit (BIO-52072).

MgCl₂: The MgCl₂ concentration in the 1x reaction mix is 3 mM. In the majority of real-time PCR conditions this is optimal for both the reverse transcriptase and the hot-start DNA polymerase.

PCR controls: It is important to detect the presence of contaminating DNA that may affect the reliability of the data. Always include a no-template control (NTC) reaction, replacing the template with PCR grade water. When performing a two-step RT-PCR, set up a no-RT control as well as an NTC for the PCR.

Optional ROX: The SensiFAST SYBR® Hi-ROX Kit is premixed with ROX (5-carboxy-X-rhodamine, succinimidyl ester), so that where necessary, ROX fluorescence can be optionally detected on certain real-time instruments. If your real-time instrument has the capability of using ROX and you wish to use this option, then this option must be selected by the user in the software.

Procedure

Reaction mix composition: Prepare a PCR mastermix. The volumes given below are based on a standard 20 µL final reaction mix and can be scaled accordingly.

Reagent	Volume	Final concentration
2x SensiFAST SYBR [®] Hi-ROX Mix	10 µL	1x
10 µM forward primer	0.8 µL	400 nM
10 µM reverse primer	0.8 µL	400 nM
Template	up to 8.4 µL	
H ₂ O	As required	
20 µL Final volume		

Sensitivity testing and C_t values: When comparing SensiFAST with a mix from another supplier we strongly recommend amplifying from a 10-fold template dilution series. Loss of detection at low template concentration is the only direct measurement of sensitivity. An early C_t value is not an indication of good sensitivity, but rather an indication of speed. In some instances increasing final MgCl₂ concentration to 6mM will reduce C_ts for difficult amplicons.

Suggested real-time PCR conditions: The following real-time PCR conditions are suitable for the SensiFAST SYBR[®] Hi-ROX Kit with the amplicons of up to 200 bp. However, the cycling conditions can be varied to suit different machine-specific protocols. It is not recommended to use annealing temperatures below 60 °C or combined annealing/extension times longer than 30 seconds.

Troubleshooting guide

Problem	Possible Cause	Recommendation
No amplification trace AND No product on agarose gel	Activation time too short	For cDNA templates, make sure SensiFAST SYBR [®] Hi-ROX is activated for 2 min at 95 °C before cycling. For more complex templates such as genomic DNA, increase activation time up to 3 minutes.
	Error in protocol setup	Verify that correct reagent concentrations, volumes, dilutions and storage conditions have been used
	Suboptimal primer design	Use primer design software or validated primers. Test primers on a control template
	Incorrect concentration of primers	Use primer concentration between 100 nM and 1 µM
	Template degraded	Re-isolate your template from the sample material or use freshly prepared template dilution
	Primers degraded	Use newly synthesized primers
	Template contaminated with PCR inhibitors	Further dilute template before PCR or purify template and resuspend it in PCR-grade H ₂ O
	Template concentration too low	Increase concentration used
Cycling conditions not optimal	Increase extension/annealing times, increase cycle number	

SensiFAST SYBR[®] Hi-ROX Kit is compatible with either three-step or two-step cycling:

3-step cycling

Cycles	Temp.	Time	Notes
1	*95 °C	*2 min	Polymerase activation
40	95 °C 60-65 °C 72 °C	5 s 10 s **5-20 s	Denaturation Annealing Extension (acquire at end of step)

*2 min for cDNA, 3 min for genomic DNA
**Not recommended to extend beyond 20 seconds

2-step cycling

Cycles	Temp.	Time	Notes
1	*95 °C	*2 min	Polymerase activation
40	95 °C 60-65 °C	5 s **15-30 s	Denaturation Annealing/extension (acquire at end of step)

*2 min for cDNA, 3 min for genomic DNA
**Not recommended to anneal/extend beyond 30 seconds

Optional analysis: After the reaction has reached completion, refer to the instrument instructions for the option of melt-profile analysis.

Troubleshooting guide (Continued)

Problem	Possible Cause	Recommendation
No amplification trace AND PCR product present on agarose gel	Error in instrument setup	Check that the acquisition settings are correct during cycling
Non-specific amplification product AND Primer-dimers	Suboptimal primer design	Redesign primers using appropriate software or use validated primers
	Primer concentration too high	Test dilution series of primer concentrations until primer dimer/non-specific amplification products disappear
	Primer concentration too low	Titrate primers in the concentration range of 100 nM - 1 µM
	Primer annealing/extension temperature(s) too low	Due to the high ionic strength of SensiFAST SYBR [®] Lo-ROX Kit it is not recommended to use annealing/extension temperatures below 60 °C. Annealing/extension temperature can be increased in steps of 2 °C in the event of non-specific products
	Template concentration too low	Increase template concentration
	Template concentration too high	Reduce template concentration until non-specific products disappear
Variability between replicates	Extension time too long	Reduce extension time to determine whether non-specific products are reduced
	Error in reaction set-up	Prepare large volume mastermix, vortex thoroughly and aliquot into reaction plate
Late amplification trace	Air bubbles in reaction mix	Centrifuge reaction samples/plate prior to running on a real-time instrument
	Activation time too short	Ensure the reaction is activated for between 1 min and 3 min at 95 °C before cycling
	Extension time too short	Increasing the extension time may be necessary for amplification products over 200 bp; double extension time to determine whether the cycle threshold (C _t) is affected
	Annealing temperature too high	Decrease annealing temperature in steps of 2 °C
	Template concentration too low	Increase concentration if possible
	Template with high secondary structure	Increase reverse transcription reaction time up to 30 min Increase reverse transcription reaction temperature up to 45 °C
	Template is degraded	Re-isolate template from sample material or use freshly prepared template dilution
	Suboptimal design of primers	Redesign primers using appropriate software or use validated primers
PCR efficiency below 90%	Primer concentration too low	Increase concentration of primer in 100 nM increments
	MgCl ₂ concentration insufficient	Increase final MgCl ₂ concentration to 6 mM
	Suboptimal design of primers	Redesign primers using appropriate software or use validated primers
PCR efficiency above 110%	Extension time is too short	Increase extension time
	Template is degraded or contains PCR inhibitors	Re-isolate template from sample material or use freshly prepared template dilution or purify template and resuspend it in water
PCR efficiency above 110%	Primer concentration too low	Increase concentration of primer in 100 nM increments
	Non-specific amplification and/or primer-dimers	Use melt analysis and 4% agarose gel electrophoresis to confirm presence of non-specific amplification products. See above for preventing/removing non-specific products