



Certificate of Analysis

COA No: CA CHM-0192

Version: 01

Reverse Transcriptase

For Research Use Only

Storage Conditions:	-20°C
Lot number:	TRT-515107
Expiry date:	August 2017

Quality Control Parameters

Analysis	Specification	Result
Functional	Fragments of sizes 1.2Kb and 6.5Kb were reverse transcribed, using standard conditions. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
Endonuclease contamination	Super coiled DNA plasmid was incubated with the reverse transcriptase for 1 hour at 37°C, the absence of nicking and cutting is shown by agarose gel electrophoresis.	Passed
DNase and RNase contamination	A DNA and RNA fragment were Incubated with the reverse transcriptase for 1 hour at 37°C. < 1% degradation was observed.	Passed

Authorised by Jade James

Europe
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

Europe
Germany

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

America

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763



Certificate of Analysis

COA No: CA BB-0003

Version: 01

RT Buffer

For Research Use Only

Storage Conditions: -20°C

Lot number: TRTB-515107

Expiry date: August 2017

Quality Control Parameters

Analysis	Specification	Result
Functional	Fragment of size 1Kb was reverse transcribed with Bioscript, with a template dilution series, using standard conditions. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection 9.7×10^{-3} ng/ μ l RNase.	Passed

Authorised by Jade James

Europe
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

Europe
Germany

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

America

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763