



A Meridian Life Science® Company

PO Box 122
Alexandria, NSW 1435
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

CERTIFICATE OF ANALYSIS

Catalog # BIO-38044
Product Name: RNA Extraction Control 560
Product Description: 100 reactions RNA Extraction Control 560 Kit
Batch #: REC560-011151

QC Methods: Functional test

RNA Extraction Control 560 kit was functionally tested based on the recommended procedures listed on the product package insert. The Internal Control RNA was analyzed on the LightCycler® 480 Real-Time PCR System. RT-PCR amplification was detected on excitation 523nm and emission 568nm configuration.

Specifications:

Analysis	Results
RNA functional test	Pass

Result

Product meets all specification

This product is for research use only, not intended for diagnostic use

Released by:

Title:

Date:



Technical Services Officer

Trademarks: LightCycler® is trademark of Roche.