



A Meridian Life Science® Company

PO Box 122  
Alexandria, NSW 1435  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763

## CERTIFICATE OF ANALYSIS

Catalog #                      BIO-35028  
Product Name:                DNA Extraction Control 670  
Product Description:        500 reactions DNA Extraction Control 670 Kit  
Batch #:                        DEC670-05216B

### QC Methods: Functional test

DNA Extraction Control 670 kit was tested based on the recommended procedures listed on the product package insert. The Internal Control DNA was analyzed on the LightCycler® 480 Real-Time PCR System. Amplification was detected on excitation 615nm and emission 670nm configuration.

### Specifications:

Analysis	Results
Functional test	Pass

### Result

Product meets all specification

This product is for research use only, not intended for diagnostic use

Released by: \_\_\_\_\_

Title: Senior Scientist

Date: 9/02/2016

**Trademarks:** LightCycler® is trademark of Roche.