BIOLINE A Meridian Life Science Company	**Certificate of Analysis**	**COA No: CA BNN-0002** **Version: 03**

dNTP Mix 40mM For Research Use Only	Storage Conditions:	-20°C
	Lot number:	DM4-516104
	Expiry date:	May 2018

Quality Control Parameters

Analysis	Specification	Result
Functional	A 3Kb human Lambda DNA fragment is amplified with a dilution series of the dNTP mix, using standard conditions and 30 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of *E. coli* and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5 x 10^{-3} U DNase.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection 9.7x10^{-3} ng/µL RNase.	Passed
Nicking Activity	Incubation of dNTP Mix with supercoiled control plasmid. Analysed by agarose gel electrophoresis. Test sample does not show an increase of linearized or relaxed plasmid.	Passed

Authorised by Christopher Weatherall

United Kingdom
Headquarters UK
info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA
info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

Germany
info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

France
Info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia
info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Singapore
Info.sg@bioline.com
Tel: +65 6774 7196
Fax +65 6774 6441