



## Certificate of Analysis

COA No: CA\_BSM-0046

Version: 03

### SensiFast™ HRM Kit

For Research Use Only

Storage Conditions: -20°C

Lot number: SFHM-516406

Expiry date: July 2018

### Quality Control Parameters

Analysis	Specification	Result
Functional Melt Analysis	HRM cycling with melt analysis $\pm 0.15^\circ\text{C}$ . Genotypes called with $\geq 90\%$ confidence.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase I.	Passed

Authorised by Christopher Weatherall

**United Kingdom**  
Headquarters UK

[info.uk@bioline.com](mailto:info.uk@bioline.com)  
Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**USA**

[info.us@bioline.com](mailto:info.us@bioline.com)  
Tel: +1 508 880 8990  
Fax: +1 508 880 8993

**Germany**

[info.de@bioline.com](mailto:info.de@bioline.com)  
Tel: +49 (0)3371 681 229  
Fax: +49 (0)3371 681 244

**France**

[info.fr@bioline.com](mailto:info.fr@bioline.com)  
Tel: +33 (0)1 42 56 04 40  
Fax: +33 (0)9 70 06 62 10

**Australia**

[info.aust@bioline.com](mailto:info.aust@bioline.com)  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763

**Singapore**

[Info.sg@bioline.com](mailto:Info.sg@bioline.com)  
Tel: +65 6774 7196  
Fax +65 6774 6441