



## Certificate of Analysis

COA No: CA\_RES-0028

Version: 02

### ISOLATE II Genomic DNA Kit

For Research Use Only

Storage Conditions:	Room Temperature
Lot number:	IS504-
Expiry date:	12 months from date of purchase

### Quality Control Parameters

Component	Specification	Result	Batch
ISOLATE II Genomic DNA Spin Columns	Nucleic acid recovery >90%	Passed	SC-316I
Collection tubes	n/a	n/a	CT-316I
Lysis Buffer GL	pH = 7.9-8.1, density	Passed	GL-316I
Buffer G1	pH = 3.7-3.9, density	Passed	G1-316I
Buffer G2	density	Passed	G2-316I
Wash Buffer GW1	pH = 7.2-7.6, density	Passed	GW1-316I
Wash Buffer GW2	pH = 7.5-7.9, density, sterility	Passed	GW2-316I
Elution Buffer G	pH = 8.3-8.7, density, sterility	Passed	EBG-316I
Proteinase K	activity	Passed	PK-316I
Proteinase K Buffer	pH = 3.5-6.0, density	Passed	PKB-316I

Authorised by Christopher Weatherall

**United Kingdom**  
Headquarters UK

**USA**

**Germany**

**France**

**Australia**

**Singapore**

[info.uk@bioline.com](mailto:info.uk@bioline.com)

Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

[info.us@bioline.com](mailto:info.us@bioline.com)

Tel: +1 508 880 8990  
Fax: +1 508 880 8993

[info.de@bioline.com](mailto:info.de@bioline.com)

Tel: +49 (0)3371 681 229  
Fax: +49 (0)3371 681 244

[info.fr@bioline.com](mailto:info.fr@bioline.com)

Tel: +33 (0)1 42 56 04 40  
Fax: +33 (0)9 70 06 62 10

[info.aust@bioline.com](mailto:info.aust@bioline.com)

Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763

[info.sg@bioline.com](mailto:info.sg@bioline.com)

Tel: +65 6774 7196  
Fax: +65 6774 6441