



Certificate of Analysis

COA No: CA_BSM-0091

Version: 01

SensiFAST Lyo-Ready No-ROX Mix, 2x

For Research Use Only

Storage Conditions: -20°C

Lot number: LNR-617201

Expiry date: February 2019

Quality Control Parameters

| Analysis | Specification | Result |
|---------------------|--|--------|
| Functional | Quantitative PCR analysis amplifying 6 genes from a dilution series of mouse cDNA under standard conditions. C _t profiles must be consistent for the test and reference sample with 0.5 +/- C _q variance. | Passed |
| DNA contamination | Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in concordance with control sample. | Passed |
| DNase contamination | Incubation of a 1 Kb double stranded DNA fragment. Incubation for 4 hours at 37° C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must exhibit less degradation than the limit of detection 2.5 x 10 ⁻³ U DNase. | Passed |

Authorised by Christopher Weatherall

United Kingdom
Headquarters UK

USA

Germany

France

Australia

Singapore

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

info.sg@bioline.com
Tel: +65 6774 7196
Fax: +65 6774 6441