



Certificate of Analysis

COA No: CA_BDB-0024

Version: 01

JetSeq Dilution Buffer

For Research Use Only

Storage Conditions: -20°C

Lot number: JDB-717103

Expiry date: April 2019

Quality Control Parameters

Analysis	Specification	Result
Functional	The DNA melting property of the JetSeq Dilution Buffer was controlled by qPCR under standard JetSeq Quantification kit conditions. The resulting melting profile should show only one major melting peak with an expected T_m value of 82.3 ± 0.4 °C.	Passed
DNase contamination	The effect of the incubation of JetSeq Dilution Buffer (4h, 37 °C) with a 1 Kb dsDNA fragment is compared with a dilution series of DNase I on agarose gel electrophoresis. Test sample must exhibit less degradation than 2.5×10^{-3} U DNase.	Passed

Authorised by Christopher Weatherall

United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

Germany

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

France

info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Singapore

Info.sg@bioline.com
Tel: +65 6774 7196
Fax +65 6774 6441



Certificate of Analysis

COA No: CA_BMM-0028

Version: 01

JetSeq Primer Mix

For Research Use Only

Storage Conditions: -20°C

Lot number: JPM-717303

Expiry date: April 2019

Quality Control Parameters

Analysis	Specification	Result
Functional	JetSeq Primer mix is used in qPCR under standard JetSeq Library Quantification kit conditions to amplify a reference DNA template. The amplification curve analysis should demonstrate an average Ct value of 9.3 ± 0.5 and the melt curve analysis is expected to produce a single peak with a T_m value of 82.3 ± 0.4 °C.	Passed
DNase contamination	The effect of the incubation of JetSeq Primer Mix (4h, 37 °C) with a 1 Kb dsDNA fragment is compared with a dilution series of DNase I on agarose gel electrophoresis. Test sample must exhibit less degradation than 2.5×10^{-3} U DNase.	Passed

Authorised by Christopher Weatherall

United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

Germany

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

France

info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Singapore

[Info.sg@bioline.com](mailto:info.sg@bioline.com)
Tel: +65 6774 7196
Fax +65 6774 6441



Certificate of Analysis

COA No: CA_BH-0055, -0056, -0057, -0058, -0059 & -0060

Version: 01

JetSeq Standards

For Research Use Only

Storage Conditions: -20°C

Lot number: JS01-717203
JS02-717203
JS03-717203
JS04-717203
JS05-717203
JS06-717203

Expiry date: April 2019

Quality Control Parameters

Analysis	Specification	Result
Functional	JetSeq Standards are used in qPCR under JetSeq Library Quantification kit recommended conditions. The average Ct value of the Standard 1 exhibits 9.3 ± 0.5 and the measured efficiency of the reaction should be between 90 – 100 %. The melt analysis should produce a single peak.	Passed
DNase contamination	The effect of the incubation of JetSeq Standards (4h, 37 °C) with a 1 Kb dsDNA fragment is compared with a dilution series of DNase I on agarose gel electrophoresis. Test sample must exhibit less degradation than 2.5×10^{-3} U DNase.	Passed

Authorised by Christopher Weatherall

United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

Germany

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

France

info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Singapore

Info.sg@bioline.com
Tel: +65 6774 7196
Fax +65 6774 6441



Certificate of Analysis

COA No: CA_BSM-0021-3

Version: 01

JetSeq FAST Lo-ROX Mix

For Research Use Only

Storage Conditions: -20°C

Lot number: JFL-617303

Expiry date: April 2019

Quality Control Parameters

Description	Specification	Result
Functional	Quantitative PCR analysis amplifying 6 genes from a dilution series of mouse cDNA under standard conditions. Cq and melting profiles must be consistent for the test and reference sample with 0.5+/- Cq variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of E. coli and mouse genomic DNA checked. Test sample must amplify in concordance with control sample.	Passed
DNase contamination	Incubation of a 1 Kb double stranded DNA fragment. Incubation for 4 hours at 37 °C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must exhibit less degradation than the limit of detection 2.5 x 10 ³ U DNase.	Passed

Authorised by Christopher Weatherall

United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

Germany

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

France

info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Singapore

Info.sg@bioline.com
Tel: +65 6774 7196
Fax +65 6774 6441