



A Meridian Life Science® Company

PO Box 122
Alexandria, NSW 1435
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

CERTIFICATE OF ANALYSIS

Catalog # BIO-35028
Product Name: DNA Extraction Control 670
Product Description: 500 reactions DNA Extraction Control 670 Kit
Batch #: CT557-BO09150
Expiry date: June 2018

QC Methods: Functional test

DNA Extraction Control 670 kit was tested based on the recommended procedures listed on the product package insert. The Internal Control DNA was analyzed on the LightCycler® 480 II Real-Time PCR System. Amplification was detected on excitation 618nm and emission 660nm configuration.

Specifications:

Analysis	Results
Functional test	Pass

Result


Product meets all specification

This product is for research use only, not intended for diagnostic use

Released by:

Title:

Date:


QA/RA Senior Scientist

10/04/2017

Trademarks: LightCycler® is trademark of Roche.