



## Certificate of Analysis

COA No: CA BH-0002

Version: 02

# HyperLadder™ 50bp

For Research Use Only

Storage Conditions: -20°C

Lot number: H2-617106

Expiry date: July 2019

### Quality Control Parameters

Analysis	Specification	Result
Functional	Single distinct bands were observed, sized 2000, 1800, 1600, 1400, 1200, 1000, 800, 700, 600, 500, 400, 300, 200, 100 and 50 bp, with agarose gel electrophoresis (ethidium stained).	Passed

Authorised by Christopher Weatherall

**United Kingdom**  
Headquarters UK

[info.uk@bioline.com](mailto:info.uk@bioline.com)  
Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**USA**

[info.us@bioline.com](mailto:info.us@bioline.com)  
Tel: +1 508 880 8990  
Fax: +1 508 880 8993

**Germany**

[info.de@bioline.com](mailto:info.de@bioline.com)  
Tel: +49 (0)3371 681 229  
Fax: +49 (0)3371 681 244

**France**

[Info.fr@bioline.com](mailto:Info.fr@bioline.com)  
Tel: +33 (0)1 42 56 04 40  
Fax: +33 (0)9 70 06 62 10

**Australia**

[info.aust@bioline.com](mailto:info.aust@bioline.com)  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763

**Singapore**

[Info.sg@bioline.com](mailto:Info.sg@bioline.com)  
Tel: +65 6774 7196  
Fax: +65 6774 6441



## Certificate of Analysis

COA No: CA BB-0005

Version: 02

### DNA Loading Buffer Blue

For Research Use Only

Storage Conditions: -20°C

Lot number: HLBB-617106

Expiry date: July 2019

### Quality Control Parameters

Analysis	Specification	Result
Functional	Tested on a 1.5% gel with 4 different sized DNA. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase.	Passed

Authorised by Christopher Weatherall

**United Kingdom**  
Headquarters UK

**USA**

**Germany**

**France**

**Australia**

**Singapore**

[info.uk@bioline.com](mailto:info.uk@bioline.com)

Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

[info.us@bioline.com](mailto:info.us@bioline.com)

Tel: +1 508 880 8990  
Fax: +1 508 880 8993

[info.de@bioline.com](mailto:info.de@bioline.com)

Tel: +49 (0)3371 681 229  
Fax: +49 (0)3371 681 244

[info.fr@bioline.com](mailto:info.fr@bioline.com)

Tel: +33 (0)1 42 56 04 40  
Fax: +33 (0)9 70 06 62 10

[info.aust@bioline.com](mailto:info.aust@bioline.com)

Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763

[info.sg@bioline.com](mailto:info.sg@bioline.com)

Tel: +65 6774 7196  
Fax: +65 6774 6441