

 <small>A Merck Life Science Company</small>	<h2>Certificate of Analysis</h2>	COA No: CA_RES-0026
		Version: 03

<h1>ISOLATE II Genomic DNA Kit</h1> For Research Use Only	Storage Conditions:	Room Temperature
	Lot number:	IS502-BO50620
	Expiry date:	12 months from date of purchase

Quality Control Parameters

Component	Specification	Result	Batch
ISOLATE II Genomic DNA Spin Columns	Nucleic acid recovery >90%	Passed	SPI0022075
Collection tubes	n/a	n/a	COL001550S
Lysis Buffer GL	pH = 7.9-8.1, density	Passed	74044
Buffer G3	density	Passed	74053
Wash Buffer GW1	pH = 7.2-7.6, density	Passed	74041
Wash Buffer GW2	pH = 7.5-7.9, density, sterility	Passed	74259
Elution Buffer G	pH = 8.3-8.7, density, sterility	Passed	74023
Proteinase K	activity	Passed	P17066S
Proteinase K Buffer	pH = 3.5-6.0, density	Passed	74050

Authorised by Christopher Weatherall



United Kingdom
Headquarters UK

USA

Germany

France

Australia

Singapore

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

info.sg@bioline.com
Tel: +65 6774 7196
Fax: +65 6774 6441