

 <small>A Merck Life Science Company</small>	<h2>Certificate of Analysis</h2>	COA No: CA_RES-0027
		Version: 03

<h1>ISOLATE II Genomic DNA Kit</h1> For Research Use Only	Storage Conditions:	Room Temperature
	Lot number:	IS503-BO51890
	Expiry date:	12 months from date of purchase

Quality Control Parameters

Component	Specification	Result	Batch
ISOLATE II Genomic DNA Spin Columns	Nucleic acid recovery >90%	Passed	SP1002243S
Collection tubes	n/a	n/a	COL001603S
Lysis Buffer GL	pH = 7.9-8.1, density	Passed	74725
Buffer G3	density	Passed	74712
Wash Buffer GW1	pH = 7.2-7.6, density	Passed	74709
Wash Buffer GW2	pH = 7.5-7.9, density, sterility	Passed	74717
Elution Buffer G	pH = 8.3-8.7, density, sterility	Passed	74713
Proteinase K	activity	Passed	P17090S
Proteinase K Buffer	pH = 3.5-6.0, density	Passed	74751

Authorised by Christopher Weatherall



United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

Germany

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

France

info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Singapore

info.sg@bioline.com
Tel: +65 6774 7196
Fax: +65 6774 6441