



A Meridian Life Science® Company

PO Box 122  
Alexandria, NSW 1435  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763

## CERTIFICATE OF ANALYSIS

Catalog # BIO-35028  
Product Name: DNA Extraction Control 670  
Product Description: 500 reactions DNA Extraction Control 670 Kit  
Batch #: CT557-BO10640  
Expiry date: JANUARY 2019

### QC Methods: Functional test

DNA Extraction Control 670 kit was tested based on the recommended procedures listed on the product package insert. The Internal Control DNA was analyzed on the LightCycler® 480 II Real-Time PCR System. Amplification was detected on excitation 618nm and emission 660nm configuration.

### Specifications:

Analysis	Results
Functional test	Pass

### Result


Product meets all specification

This product is for research use only, not intended for diagnostic use

Released by:

Title:

Date:

  
QA/RA Senior Scientist

14/09/2017

**Trademarks:** LightCycler® is trademark of Roche.