

 <small>A Merck Life Science Company</small>	<h2>Certificate of Analysis</h2>	COA No: CA_RES-0026
		Version: 03

<h1>ISOLATE II Genomic DNA Kit</h1>  For Research Use Only	Storage Conditions:	Room Temperature
	Lot number:	IS502-BO54750
	Expiry date:	12 months from date of purchase

### Quality Control Parameters

Component	Specification	Result	Batch
ISOLATE II Genomic DNA Spin Columns	Nucleic acid recovery >90%	Passed	SPI002278S
Collection tubes	n/a	n/a	COL001613S
Lysis Buffer GL	pH = 7.9-8.1, density	Passed	76355
Buffer G3	density	Passed	76349
Wash Buffer GW1	pH = 7.2-7.6, density	Passed	76374
Wash Buffer GW2	pH = 7.5-7.9, density, sterility	Passed	76342
Elution Buffer G	pH = 8.3-8.7, density, sterility	Passed	76297
Proteinase K	activity	Passed	P17116
Proteinase K Buffer	pH = 3.5-6.0, density	Passed	76287

Authorised by Christopher Weatherall



**United Kingdom**  
Headquarters UK

[info.uk@bioline.com](mailto:info.uk@bioline.com)  
Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**USA**

[info.us@bioline.com](mailto:info.us@bioline.com)  
Tel: +1 508 880 8990  
Fax: +1 508 880 8993

**Germany**

[info.de@bioline.com](mailto:info.de@bioline.com)  
Tel: +49 (0)3371 681 229  
Fax: +49 (0)3371 681 244

**France**

[info.fr@bioline.com](mailto:info.fr@bioline.com)  
Tel: +33 (0)1 42 56 04 40  
Fax: +33 (0)9 70 06 62 10

**Australia**

[info.aust@bioline.com](mailto:info.aust@bioline.com)  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763

**Singapore**

[info.sg@bioline.com](mailto:info.sg@bioline.com)  
Tel: +65 6774 7196  
Fax: +65 6774 6441