



Certificate of Analysis

COA No: CA_RES-0033

Version: 03

ISOLATE II RNA Mini Kit

For Research Use Only

Storage Conditions:	Room Temperature
Lot number:	IS509-BO55860
Expiry date:	12 months from date of purchase

Quality Control Parameters

Component	Specification	Result	Batch
ISOLATE II Filters	n/a	n/a	SP1002294
ISOLATE II RNA Mini Columns	Nucleic acid recovery >90%	Passed	SPB002219
Collection tubes (2ml)	n/a	n/a	COL001641
Collection tubes (1.5ml)	n/a	n/a	051017
Lysis Buffer RLY	pH = 6.7-7.3, density	Passed	767030
Wash Buffer RW1	pH = 6.9-7.5, density	Passed	76715
Wash Buffer RW2	pH = 7.5-7.9, density, sterility	Passed	76715
Membrane Desalting Buffer MEM	pH = 7.8-7.1, density	Passed	76700
Reaction Buffer for DNase I RDN	pH = 6.9-7.1, density	Passed	76739
DNase, RNase-free	No RNase activity detectable in connection with our kit	Passed	D1716R
RNase-free Water	Density, no RNase activity, sterility	Passed	76740

Authorised by Ivan Mijatovic

United Kingdom
Headquarters UK

USA

Germany

France

Australia

Singapore

info.uk@bioline.com

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

info.us@bioline.com

Tel: +1 508 880 8990
Fax: +1 508 880 8993

info.de@bioline.com

Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

info.fr@bioline.com

Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

info.aust@bioline.com

Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

info.sg@bioline.com

Tel: +65 6774 7196
Fax: +65 6774 6441