



Certificate of Analysis

COA No: CA_XBH-0005

Version: 04

HyperLadder™ 25bp

For Research Use Only

Storage Conditions: -20°C

Lot number: H5-718106A

Expiry date: July 2020

Quality Control Parameters

Certified Values:

| Number of Bases | Method of Testing | Specification | Method of Testing | Results |
|-----------------|-------------------|-------------------|---|---------|
| 25 bp | Sequencing | 80 ng/band ± 10% | UV absorption spectrum Visual comparison test vs history | Passed |
| 50 bp | Sequencing | 80ng/band ± 10% | UV absorption spectrum Visual comparison test vs history | Passed |
| 75 bp | Sequencing | 80 ng/band ± 10% | UV absorption spectrum Visual comparison test vs history | Passed |
| 100 bp | Sequencing | 120 ng/band ± 10% | UV absorption spectrum Visual comparison test vs history | Passed |
| 125 bp | Sequencing | 80 ng/band ± 10% | UV absorption spectrum Visual comparison test vs history | Passed |
| 150 bp | Sequencing | 80 ng/band ± 10% | UV absorption spectrum Visual comparison test vs history | Passed |
| 175 bp | Sequencing | 80 ng/band ± 10% | UV absorption spectrum Visual comparison test vs history | Passed |
| 200 bp | Sequencing | 120 ng/band ± 10% | UV absorption spectrum Visual comparison test vs history | Passed |
| 250 bp | Sequencing | 60 ng/band ± 10% | UV absorption spectrum Visual comparison test vs history | Passed |
| 300 bp | Sequencing | 60 ng/band ± 10% | UV absorption spectrum Visual comparison test vs history | Passed |
| 400 bp | Sequencing | 60 ng/band ± 10% | UV absorption spectrum Visual comparison test vs history | Passed |

United Kingdom
Headquarters UK

USA

Germany

France

Australia

Singapore

info.uk@bioline.com

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

info.us@bioline.com

Tel: +1 508 880 8990
Fax: +1 508 880 8993

info.de@bioline.com

Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

Info.fr@bioline.com


Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

info.aust@bioline.com

Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Info.sg@bioline.com

Tel: +65 6774 7196
Fax: +65 6774 6441

| | | |
|--|----------------------------------|---------------------|
|  | <h2>Certificate of Analysis</h2> | COA No: CA_XBH-0005 |
| | | Version: 04 |

| | | | | |
|--------|------------|----------------------|---|--------|
| 500 bp | Sequencing | 60 ng/band \pm 10% | UV absorption spectrum Visual comparison test vs history | Passed |
|--------|------------|----------------------|---|--------|

Note: The values given relate to individual bands. Following the combination of all bands in one solution, the Ladder may be used for approximating the mass of DNA.

Authorised by Ivan Mijatovic



United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

Germany

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

France

info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Singapore

Info.sg@bioline.com
Tel: +65 6774 7196
Fax +65 6774 6441

| | | |
|---|--------------------------------|---------------------|
|  <small>A Merck Life Science Company</small> | Certificate of Analysis | COA No: CA XBB-0005 |
| | | Version: 03 |

| | | |
|--|---------------------|--------------|
| <h2>DNA Loading Buffer Blue</h2> <p>For Research Use Only</p> | Storage Conditions: | -20°C |
| | Lot number: | HLBB-618106B |
| | Expiry date: | July 2020 |

| |
|-----------------------------------|
| Quality Control Parameters |
|-----------------------------------|

| Analysis | Specification | Result |
|---------------------|---|--------|
| Functional | Tested on a 1.5% gel with 4 different sized DNA. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained). | Passed |
| DNase contamination | Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase. | Passed |

Authorised by Ivan Mijatovic



United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

Germany

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

France

Info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Singapore

Info.sg@bioline.com
Tel: +65 6774 7196
Fax: +65 6774 6441