



A Meridian Life Science® Company

PO Box 122  
Alexandria, NSW 1435  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763

## CERTIFICATE OF ANALYSIS

Catalog #                    BIO-35029  
Product Name:              DNA Extraction Control 670  
Product Description:      2000 reactions DNA Extraction Control 670 Kit  
Batch #:                    CT558-BO12420  
Expiry date:                JULY 2019

### QC Methods: Functional test

DNA Extraction Control 670 kit was tested based on the recommended procedures listed on the product package insert. The Internal Control DNA was analyzed on the LightCycler® 480 II Real-Time PCR System. Amplification was detected on excitation 618nm and emission 660nm configuration.

### Specifications:

Analysis	Results
Functional test	Pass

### Result

Product meets all specification

This product is for research use only, not intended for diagnostic use

Released by: \_\_\_\_\_

Title: \_\_\_\_\_

Date: \_\_\_\_\_

QA/RA Senior Scientist

1/06/2018

**Trademarks:** LightCycler® is trademark of Roche.