

 <small>A Molecular Life Sciences Company</small>	<h2>Certificate of Analysis</h2>	COA No: CA_XBS-0085
		Version: 03

<h1>Oligo (dT)<sub>18</sub></h1> <p>For Research Use Only</p>	Storage Conditions:	-20°C
	Lot number:	ODT-718107B
	Expiry date:	August 2020

### Quality Control Parameters

Analysis	Specification	Result
Functional	A 1Kb fragment is reverse transcribed from Poly A RNA with a dilution series of Bioscript, using standard conditions. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection $9.7 \times 10^{-3}$ ng/ $\mu$ L RNase.	Passed

Authorised by Ivan Mijatovic



**United Kingdom**  
Headquarters UK

[info.uk@bioline.com](mailto:info.uk@bioline.com)  
Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**USA**

[info.us@bioline.com](mailto:info.us@bioline.com)  
Tel: +1 508 880 8990  
Fax: +1 508 880 8993

**Germany**

[info.de@bioline.com](mailto:info.de@bioline.com)  
Tel: +49 (0)3371 681 229  
Fax: +49 (0)3371 681 244

**France**

[info.fr@bioline.com](mailto:info.fr@bioline.com)  
Tel: +33 (0)1 42 56 04 40  
Fax: +33 (0)9 70 06 62 10

**Australia**

[info.aust@bioline.com](mailto:info.aust@bioline.com)  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763

**Singapore**

[Info.sg@bioline.com](mailto:Info.sg@bioline.com)  
Tel: +65 6774 7196  
Fax +65 6774 6441