



Certificate of Analysis

COA No: CA_RES-0028

Version: 04

ISOLATE II Genomic DNA Kit

For Research Use Only

Storage Conditions:	Room Temperature
Lot number:	IS504-BO60960
Expiry date:	12 months from date of purchase

Quality Control Parameters

Component	Specification	Result	Batch
ISOLATE II Genomic DNA Spin Columns	Nucleic acid recovery >90%	Passed	SPI002278S
Collection tubes	n/a	n/a	COL001614S
Lysis Buffer GL	pH = 7.9-8.1, density	Passed	79306
Buffer G3	density	Passed	79285
Wash Buffer GW1	pH = 7.2-7.6, density	Passed	79299
Wash Buffer GW2	pH = 7.5-7.9, density, sterility	Passed	79304
Elution Buffer G	pH = 8.3-8.7, density, sterility	Passed	79305
Proteinase K	activity	Passed	P18041S
Proteinase K Buffer	pH = 3.5-6.0, density	Passed	79294

Authorised by Ivan Mijatovic

United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

Germany

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

France

info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Singapore

Info.sg@bioline.com
Tel: +65 6774 7196
Fax: +65 6774 6441