



## Certificate of Analysis

COA No: CA BMM-0025

Version: 03

# MyTaq Blood PCR Mix

For Research Use Only

Storage Conditions:	-20°C
Lot number:	MTBX-818208A
Expiry date:	September 2020

### Quality Control Parameters

Analysis	Specification	Result
Functional	<p>Fragment of size 5Kb was amplified with a dilution series human blood, using standard conditions and 40 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).</p> <p>A four plex PCR reaction was run using standard conditions, 40 cycles with a dilution series of mix and 35 cycles with a dilution series of human blood. Bands sizes were 1.2Kb, 844bp, 450bp and 237bp. Distinct bands were observed with agarose gel electrophoresis (ethidium stained).</p>	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase.	Passed

Authorised by Ivan Mijatovic

**United Kingdom**  
Headquarters UK

**USA**

**Germany**

**France**

**Australia**

**Singapore**

[info.uk@bioline.com](mailto:info.uk@bioline.com)  
Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

[info.us@bioline.com](mailto:info.us@bioline.com)  
Tel: +1 508 880 8990  
Fax: +1 508 880 8993

[info.de@bioline.com](mailto:info.de@bioline.com)  
Tel: +49 (0)3371 681 229  
Fax: +49 (0)3371 681 244

[info.fr@bioline.com](mailto:info.fr@bioline.com)  
Tel: +33 (0)1 42 56 04 40  
Fax: +33 (0)9 70 06 62 10

[info.aust@bioline.com](mailto:info.aust@bioline.com)  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763

[Info.sg@bioline.com](mailto:Info.sg@bioline.com)  
Tel: +65 6774 7196  
Fax +65 6774 6441