



Certificate of Analysis

COA No: CA_XBH-0002

Version: 05

HyperLadder™ 50bp

For Research Use Only

Kit Lot No:	BIO-33054_MW424-B066230
Storage Conditions:	-20°C
Component Lot No:	H2-819101A
Expiry date:	February 2021

Quality Control Parameters

Certified Values:

Number of Bases	Method of Testing	Specification	Method of Testing	Results
50 bp	Sequencing	40 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
100 bp	Sequencing	40 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
200 bp	Sequencing	40 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
300 bp	Sequencing	100 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
400 bp	Sequencing	30 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
500 bp	Sequencing	30 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
600 bp	Sequencing	30 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
700 bp	Sequencing	30 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
800 bp	Sequencing	30 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
1000 bp	Sequencing	100 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
1200 bp	Sequencing	20 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
1400 bp	Sequencing	20 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed

United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info@meridianlifescience.com
Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

info.de@bioline.com
Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01

France

info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763



Certificate of Analysis

COA No: CA_XBH-0002

Version: 05

1600 bp	Sequencing	20 ng/band \pm 10%	UV absorption spectrum Visual comparison test vs history	Passed
1800 bp	Sequencing	20 ng/band \pm 10%	UV absorption spectrum Visual comparison test vs history	Passed
2000 bp	Sequencing	50 ng/band \pm 10%	UV absorption spectrum Visual comparison test vs history	Passed

Note: The values given relate to individual bands. Following the combination of all bands in one solution, the Ladder may be used for approximating the mass of DNA.

Authorised by Christopher Weatherall

United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info@meridianlifescience.com
Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

info.de@bioline.com
Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01

France

info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763



Certificate of Analysis

COA No: CA XBB-0005

Version: 04

DNA Loading Buffer Blue

For Research Use Only

Kit Lot No:	BIO-33054_MW424-B066230
Storage Conditions:	-20°C
Component Lot No:	HLBB-2035.005
Expiry date:	February 2021

Quality Control Parameters

Analysis	Specification	Result
Functional	Tested on a 1.5% gel with 4 different sized DNA. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase.	Passed

Authorised by Christopher Weatherall

United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info@meridianlifescience.com
Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

info.de@bioline.com
Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01

France

info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763