



Certificate of Analysis

COA No: CA_RES-0021

Version: 04

ISOLATE II Gel and PCR Kit

For Research Use Only

Kit Lot No:	BIO-52059_IS496-B068030
Storage Conditions:	Room Temperature
Component Lot No:	IS496-B068030
Expiry date:	12 months from date of purchase

Quality Control Parameters

Component	Specification	Result	Batch
ISOLATE II PCR and Gel Columns	Nucleic acid recovery >90%	Passed	SPI002456
Collection tubes	n/a	n/a	COL001720
Binding Buffer CB	pH = 4.8-5.2, density	Passed	82776
Wash Buffer CW (conc)	pH = 7.5-7.9, density	Passed	82751
Elution Buffer C	pH = 8.3-8.7, density	Passed	82691

Authorised by Christopher Weatherall

United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info@meridianlifescience.com
Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

info.de@bioline.com
Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01

France

info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763