



Certificate of Analysis

COA No: CA_RES-0028

Version: 05

ISOLATE II Genomic DNA Kit

For Research Use Only

Kit Lot No:	BIO-52067_IS504-B068310
Storage Conditions:	Room Temperature
Component Lot No:	IS504-B068310
Expiry date:	12 months from date of purchase

Quality Control Parameters

Component	Specification	Result	Batch
ISOLATE II Genomic DNA Spin Columns	Nucleic acid recovery >90%	Passed	SPI002384S
Collection tubes	n/a	n/a	COL001725
Lysis Buffer GL	pH = 7.9-8.1, density	Passed	82614
Buffer G3	density	Passed	82184
Wash Buffer GW1	pH = 7.2-7.6, density	Passed	82603
Wash Buffer GW2	pH = 7.5-7.9, density, sterility	Passed	82663
Elution Buffer G	pH = 8.3-8.7, density, sterility	Passed	82624
Proteinase K	activity	Passed	P18114S
Proteinase K Buffer	pH = 3.5-6.0, density	Passed	82925

Authorised by Christopher Weatherall

United Kingdom
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

info@meridianlifescience.com
Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

info.de@bioline.com
Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01

France

info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763