

## dATP 100mM

Suitable for Research and further Manufacturing Use as an IVD component

Catalog No:	BIO-39026
Lot No:	DS404-B073400
Shipping / Storage Conditions:	-20°C
Component Lot No:	DA-819107B
Expiry date:	August 2021

### Quality Control Parameters

2'-deoxyadenosine-5'-triphosphate



MW = 514.916 g /mol

Certified <1% deoxynucleoside monophosphates and deoxynucleoside diphosphates

Characteristics	Specification	Result
Concentration (at $\lambda_{max}$ , pH 7.0, $\epsilon = 15.4 \text{ E x mmol}^{-1} \text{ x cm}^{-1}$ )	100 mM $\pm$ 5%	102.3
pH of Solution(at 20°C)	7.5 – 8.0	7.53 @ 22°C
$\lambda_{max}$ (at pH 7.0)	259 $\pm$ 1 nm	259.5 nm
A250/A260	0.78 $\pm$ 0.03	0.79
A280/A260	0.15 $\pm$ 0.02	0.13
Purity dATP (HPLC Area % at $\lambda_{max}$ )	$\geq$ 99%	99.8%
dNDP + dNMP (HPLC Area % at $\lambda_{max}$ )	<1%	Passed
Appearance	Clear colourless solution	Passed

Analysis	Specification	Result
Functional	A 3Kb Lambda DNA fragment is amplified with a dilution series of dATP, using standard conditions and 30 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase.	Passed
RNase	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection $9.7 \times 10^{-3}$ ng/ $\mu$ L RNase.	Passed
Nicking Activity	Incubation of dATP with supercoiled control plasmid. Analysed by agarose gel electrophoresis. Test sample does not show an increase of linearized or relaxed plasmid.	Passed

Authorised by Christopher Weatherall


## dCTP 100mM

Suitable for Research and further Manufacturing Use as an IVD component

Catalog No:	BIO-39026
Lot No:	DS404-B073400
Shipping / Storage Conditions:	-20°C
Component Lot No:	DC-819107B
Expiry date:	August 2021

### Quality Control Parameters

2'-deoxycytidine-5'-triphosphate



MW = 490.891 g /mol

Certified &lt;1% deoxynucleoside monophosphates and deoxynucleoside diphosphates

Characteristics	Specification	Result
Concentration (at $\lambda_{max}$ , pH 7.0, $\epsilon = 9.1 \text{ E} \times \text{mmol}^{-1} \times \text{cm}^{-1}$ )	100 mM $\pm$ 5%	102.7 mM
pH of Solution (at 20°C)	7.5 – 8.0	7.5 @ 22°C
$\lambda_{max}$ (at pH 7.0)	272 $\pm$ 1 nm	271 nm
A250/A260	0.82 $\pm$ 0.03	0.80
A280/A260	0.98 $\pm$ 0.03	0.96
Purity dCTP (HPLC Area % at $\lambda_{max}$ )	$\geq$ 99%	99.9%
dNDP + dNMP (HPLC Area % at $\lambda_{max}$ )	<1%	Passed
Appearance	Clear colourless solution	Passed

Analysis	Specification	Result
Functional	A 3Kb Lambda DNA fragment is amplified with a dilution series of dCTP, using standard conditions and 30 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase.	Passed
RNase	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection $9.7 \times 10^{-3}$ ng/ $\mu$ L RNase.	Passed
Nicking Activity	Incubation of dCTP with supercoiled control plasmid. Analysed by agarose gel electrophoresis. Test sample does not show an increase of linearized or relaxed plasmid.	Passed

Authorised by Christopher Weatherall

**United Kingdom**  
Headquarters UK

[info.uk@bioline.com](mailto:info.uk@bioline.com)  
Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**USA**

[info@meridianlifescience.com](mailto:info@meridianlifescience.com)  
Tel: +1 901.382.8716  
Fax: +1 901.382.0027

**Germany**

[info.de@bioline.com](mailto:info.de@bioline.com)  
Tel: +49 (0)3371 60222 00  
Fax: +49 (0)3371 60222 01

**France**

[info.fr@bioline.com](mailto:info.fr@bioline.com)  
Tel: +33 (0)1 42 56 04 40  
Fax: +33 (0)9 70 06 62 10

**Australia**

[info.aust@bioline.com](mailto:info.aust@bioline.com)  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763

## dGTP 100mM

Suitable for Research and further Manufacturing Use as an IVD component

Catalog No:	BIO-39026
Lot No:	DS404-B073400
Shipping / Storage Conditions:	-20°C
Component Lot No:	DG-919107B
Expiry date:	August 2021

### Quality Control Parameters

2'-deoxyguanosine-5'-triphosphate



MW = 530.916 g / mol

Certified <1% deoxynucleoside monophosphates and deoxynucleoside diphosphates

Characteristics	Specification	Result
Concentration (at $\lambda_{max}$ , pH 7.0, $\epsilon = 13.7 \text{ E x mmol}^{-1} \text{ x cm}^{-1}$ )	100 mM $\pm$ 5%	100.3 mM
pH of Solution (at 20°C)	7.5 – 8.0	7.53 @ 21.3°C
$\lambda_{max}$ (at pH 7.0)	252 $\pm$ 1 nm	252.5 nm
A250/A260	1.16 $\pm$ 0.05	1.18
A280/A260	0.66 $\pm$ 0.03	0.67
dNTP (HPLC Area % at $\lambda_{max}$ )	$\geq$ 99%	99.73%
dNDP + dNMP (HPLC Area % at $\lambda_{max}$ )	<1%	Passed
Appearance	Clear colourless solution	Passed

Analysis	Specification	Result
Functional	A 3Kb Lambda DNA fragment is amplified with a dilution series of dGTP, using standard conditions and 30 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase.	Passed
RNase	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection $9.7 \times 10^{-3}$ ng/ $\mu$ L RNase.	Passed
Nicking Activity	Incubation of dGTP with supercoiled control plasmid. Analysed by agarose gel electrophoresis. Test sample does not show an increase of linearized or relaxed plasmid.	Passed

Authorised by Christopher Weatherall


## dTTP 100mM

Suitable for Research and further Manufacturing Use as an IVD component

Catalog No:	BIO-39026
Lot No:	DS404-B073400
Shipping / Storage Conditions:	-20°C
Component Lot No:	DT-919107B
Expiry date:	August 2021

### Quality Control Parameters

2'-deoxythymidine-5'-triphosphate



MW = 505.903 g /mol

Certified <1% deoxynucleoside monophosphates and deoxynucleoside diphosphates

Characteristics	Specification	Result
Concentration (at $\lambda_{max}$ , pH 7.0, $\epsilon = 9.5 \text{ E x mmol}^{-1} \text{ x cm}^{-1}$ )	100 mM $\pm$ 5%	101.6 mM
pH of Solution (at 20°C)	7.5 – 8.0	7.51 @ 21.3°C
$\lambda_{max}$ (at pH 7.0)	267 $\pm$ 1 nm	267 nm
A250/A260	0.65 $\pm$ 0.03	0.64
A280/A260	0.73 $\pm$ 0.02	0.72
Purity dTTP (HPLC Area % at $\lambda_{max}$ )	$\geq$ 99%	99.55%
dNDP + dNMP (HPLC Area % at $\lambda_{max}$ )	<1%	Passed
Appearance	Clear colourless solution	Passed

Analysis	Specification	Result
Functional	A 3Kb Lambda DNA fragment is amplified with a dilution series of dTTP, using standard conditions and 30 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase.	Passed
RNase	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection $9.7 \times 10^{-3}$ ng/ $\mu$ L RNase.	Passed
Nicking Activity	Incubation of dTTP with supercoiled control plasmid. Analysed by agarose gel electrophoresis. Test sample does not show an increase of linearized or relaxed plasmid.	Passed

Authorised by Christopher Weatherall