bioline meridian BIOSCIENCE	Certificate of Analysis	COA No: CA_BSM-0041
		Version: 05

SensiFAST™ Probe Lo-ROX Kit Suitable for Research and further Manufacturing Use as an IVD component	Catalog No:	BIO-84050
	Lot No:	SF886-B078650
	Shipping / Storage Conditions:	-20°C
	Component Lot No:	SFPL-920501A
	Expiry date:	February 2022

Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative PCR analysis amplifying 6 genes from a dilution series of mouse cDNA under standard conditions. Cq profiles must be consistent for the test and reference sample with ± 0.5 Cq variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of *E. coli* and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
Exonuclease contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5 x 10^{-3} U DNase I.	Passed

Authorised by Christopher Weatherall

United Kingdom
Headquarters UK
info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA
info@meridianlifescience.com
Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany
info.de@bioline.com
Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01

France
Info.fr@bioline.com
Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10

Australia
info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763