

# MyTaq™ HS Mix

Shipping: On Dry/Blue ice Catalog numbers:

BIO-25045: 200 x 50 µL reactions 4 x 1.25 mL

Batch No.: See vial

BIO-25046: 1000 x 50 µL reactions 20 x 1.25 mL

Concentration: 2x

Store at -20 °C



## Storage and stability:

MyTaq HS Mix is shipped on dry/blue ice. On arrival store at -20 °C for optimum stability. Repeated freeze/thaw cycles should be avoided.

## Expiry:

When stored under the recommended conditions and handled correctly, full activity of the kit is retained until the expiry date on the outer box label.

## Safety precautions:

Please refer to the material safety data sheet for further information.

## Quality control specifications:

MyTaq HS Mix and its components are extensively tested for activity, processivity, efficiency, heat activation, sensitivity, absence of nuclease contamination and absence of nucleic acid contamination prior to release.

## Notes:

For research or further manufacturing use only.

## Trademarks:

HyperLadder and MyTaq are trademarks of Bioline Reagents Ltd.

## Description

MyTaq™ HS Mix is a ready-to-use 2x mix for fast, highly-specific, hot-start PCR. MyTaq HS Mix is powered by antibody mediated hot-start and does not possess polymerase activity during the reaction set-up, thus reducing non-specific amplification. The advanced formulation of MyTaq HS Mix allows fast cycling conditions to be used, greatly reducing the reaction time without compromising PCR specificity and yield. Thanks to its speed and high specificity MyTaq HS Mix is also highly suitable for end point multiplex PCR. MyTaq HS Mix contains all the reagents (including stabilizers) necessary for trouble-free PCR set up. The product is supplied conveniently all in one tube to reduce the number of pipetting steps and to facilitate increased efficiency, throughput and reproducibility.

## Components

	200 Reactions	1000 Reactions
MyTaq HS Mix, 2x	4 x 1.25 mL	20 x 1.25 mL

## Standard MyTaq HS Mix Protocol

The following protocol is for a standard 50 µL reaction and can be used as a starting point for reaction optimization. Please refer to the Important Considerations and PCR Optimization section.

### PCR set-up:

Template	200 ng
Primers (20 µM each)	1 µL
MyTaq HS Mix, 2x	25 µL
Water (dH <sub>2</sub> O)	up to 50 µL

### PCR cycling conditions:

Step	Temperature	Time	Cycles
Initial denaturation	95 °C	1 min	1
Denaturation	95 °C	15 s	25-35
Annealing*	User determined	15 s	
Extension*	72 °C	10 s	

\* These parameters may require optimization, please refer to the Important Considerations and PCR Optimization section if needed.

## Colony PCR Protocol

MyTaq HS Mix can be used for amplification of plasmid DNA directly from liquid cultures or from colonies on agar plates:

- From liquid culture: up to 8 µL of the overnight culture can be directly added to the final reaction mix.
- From colonies: we recommend using a sterile tip to stab the colony and resuspend it directly in the 50 µL reaction mix.

## Recommended cycling conditions for colony PCR of fragment up to 1 kb

Step	Temperature	Time	Cycles
Initial denaturation	95 °C	1 min	1
Denaturation	95 °C	15 s	25-35
Annealing*	User determined	15 s	
Extension*	72 °C	10 s	

\* These parameters may require optimization, please refer to the Important Considerations and PCR Optimization section if needed.

## Multiplex PCR Protocol

MyTaq HS Mix is suitable for multiplex PCR; adjustment of the cycling conditions may be required. As a starting point we recommend using the following conditions:

### Recommended standard cycling conditions for multiplex PCR

\* These parameters may require optimization, please refer to the Important

Step	Temperature	Time	Cycles
Initial denaturation	95 °C	2 min	1
Denaturation	95 °C	30 s	25*
Annealing/Extension*	User determined	4 min*	

Considerations and PCR Optimization section if needed.

## Important Considerations and PCR Optimization

The optimal conditions may vary from reaction to reaction and are dependent on the template/primers used.

**Primers:** Forward and reverse primers are generally used at the final concentration of 0.2-0.6 µM each. As a starting point, we recommend using a 0.4 µM final concentration (*i.e.* 20 pmol of each primer per 50 µL reaction volume). Too high a primer concentration can reduce the specificity of priming, resulting in non-specific products.

When designing primers we recommend using primer-design software such as Primer3 (<http://frodo.wi.mit.edu/primer3>) or visual OMP™ (<http://dnasoftware.com>) with monovalent and divalent cation concentrations of 10 mM and 3 mM respectively. Primers should have a melting temperature (T<sub>m</sub>) of approximately 60 °C.

**Template:** The amount of template in the reaction depends mainly on the type of DNA used. For templates with low structural complexity, such as plasmid DNA, we recommend using 50 pg-10 ng DNA per 50 µL reaction volume. For eukaryotic genomic DNA, we recommend a starting amount of 200 ng DNA per 50 µL reaction, this can be varied between 5 ng-500 ng. It is important to avoid using template resuspended in EDTA-containing solutions (e.g. TE buffer) since EDTA chelates free Mg<sup>2+</sup>.

**Initial denaturation:** The initial denaturation step is required to activate the enzyme and fully melt the template. We recommend 1 minute of initial denaturation at 95 °C, however for more complex templates such as eukaryotic genomic DNA, longer initial denaturation times of up to 3 minutes may be required.

**Denaturation:** Our protocol recommends a 15 s cycling denaturation step at 95 °C, which is also suited to GC-rich templates (>55%). For low GC content amplicons (40-45%), the denaturation step can be decreased to 5 s.

**Annealing temperature and time:** The optimal annealing temperature is dependent upon the primer sequences and is usually 2-5 °C below the lower T<sub>m</sub> of the pair. We recommend starting with a 55 °C annealing temperature and, if necessary, running a temperature gradient to determine the optimal annealing temperature. Depending on the reaction the annealing time can also be reduced to 5 s.

**Extension temperature and time:** The extension step should be performed at 72 °C. The extension time depends on the length of the amplicon and the complexity of the template. An extension time of 10 s is sufficient for amplicons under 1 kb or up to 5 kb for low complexity template such as plasmid DNA. For amplification of fragments over 1kb from high complexity template, such as eukaryotic genomic DNA, longer extension times are recommended. In order to find the fastest optimal condition, we suggest increasing the extension time up to 30 s/kb.

**Multiplexing:** When doing multiplex PCR the recommended 2-step cycling protocol may be optimized as follows:

- Annealing/extension temperature: we highly recommend initially using a temperature gradient to determine the optimal annealing temperature needed for the primer set used.
- Annealing/extension time: in most cases a 4 min annealing/extension step is largely sufficient. However in order to reduce the overall cycling time this step can be reduced down to 1 min, especially in the case of a lower number of multiplex amplicons.
- Cycling number: we recommend starting with 25 cycles and if necessary, optimizing this parameter. An excess of cycles may generate diffuse bands, too few may result in weak or no amplification.

## Troubleshooting Guide

Problem	Possible Cause	Recommendation
<b>No PCR product</b>	Missing component	- Check reaction set-up and volumes used
	Defective component	- Check the aspect and the concentrations of all components as well as the storage conditions. If necessary test each component individually in controlled reactions
	Cycling conditions not optimal	- Decrease the annealing temperature - Run a temperature gradient to determine the optimal annealing temperature - Increase the extension time, especially if amplifying a long target - Increase the number of cycles
	Difficult template	- Increase the denaturation time
<b>Smearing or Non-Specific products</b>	Excessive cycling	- Decrease the number of cycles
	Extension time too long	- Decrease the extension time
	Annealing temperature too low	- Increase the annealing temperature
	Primer concentration too high	- Decrease primer concentration
	Contamination	- Replace each component in order to find the possible source of contamination - Set up the PCR and analyze the PCR product in separate areas.

## Technical Support

If the troubleshooting guide does not solve the difficulty you are experiencing, please contact your local distributor or our Technical Support with details of reaction set-up, cycling conditions and relevant information.

Email: [mbi.tech@meridianlifescience.com](mailto:mbi.tech@meridianlifescience.com)

## Associated Products

Product Name	Pack Size	Cat No
Agarose	500 g	BIO-41025
Agarose tablets	300 g	BIO-41027
HyperLadder™ 1kb	200 Lanes	BIO-33025
SureClean Plus	1 x 5 mL	BIO-37047

Bioline Reagents Ltd  
UNITED KINGDOM

Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

Meridian Life Science Inc.  
USA

Tel: +1 901 382 8716  
Fax: +1 901 382 0027

Bioline GmbH  
GERMANY

Tel: +49 (0)3371 60222 00  
Fax: +49(0)3371 60222 01

Bioline (Aust) Pty. Ltd  
AUSTRALIA

Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763