


<h2>SensiMix™ SYBR® & Fluorescein Kit</h2> <p>For research or further manufacturing use only</p>	Catalog No:	QT615-05
	Lot No:	SM608-B130940
	Storage Conditions:	-20°C
	Component Lot No:	SMPF-324109A
	Expiry date:	October 2026

Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative PCR analysis amplifying 6 genes from a dilution series of mouse cDNA under standard conditions. Cq and melting profiles must be consistent for the test and reference sample with ± 0.5 Cq variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5 x 10 ⁻³ U DNase I.	Passed

QA / QC Representative:



J. Rahnenführer

Date: 16th September 2024

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01