

MyTaq™ Reaction Buffer

Suitable for Research and further Manufacturing Use as an IVD component

Catalog No:	BIO-37111
Lot No:	ES548-B074730
Shipping / Storage Conditions:	-20°C
Component Lot No:	MTB-919109B
Expiry date:	October 2021

Quality Control Parameters

Analysis	Specification	Result
Functional	Fragment of size 1200bp was amplified with a dilution series of human genomic DNA, using standard conditions and 35 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase.	Passed

Authorised by Christopher Weatherall