

SensiFast™ HRM Kit


For research or further manufacturing use only

Catalog No:	BIO-32005
Lot No:	SF628-B123930
Storage Conditions:	-20°C
Component Lot No:	SFHM-324301A
Expiry date:	February 2026

Quality Control Parameters

Analysis	Specification	Result
Functional Melt Analysis	HRM cycling with melt analysis $\pm 0.15^\circ\text{C}$. Genotypes called with $\geq 90\%$ confidence.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase I.	Passed

QA / QC Representative:



J. Rahnenführer

 Date: 16th January 2024

United Kingdom

Tel: +44 (0)20 8830 5300
 Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
 Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
 Fax: +49 (0)3371 60222 01