

SensiMix™ II Probe Kit


For research or further manufacturing use only

Catalog No:	BIO-83005
Lot No:	SM664-B124810
Storage Conditions:	-20°C
Component Lot No:	SM2-324301A
Expiry date:	February 2026

Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative PCR analysis amplifying 6 genes from a dilution series of mouse cDNA under standard conditions. Cq profiles must be consistent for the test and reference sample with ± 0.5 Cq variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5 x 10 ⁻³ U DNase I.	Passed

QA / QC Representative:



Jan Rahnenführer

 Date: 30th January 2024

United Kingdom

Tel: +44 (0)20 8830 5300
 Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
 Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
 Fax: +49 (0)3371 60222 01

ROX Solution, 25 µM


For research or further manufacturing use only

Catalog No:	BIO-83005
Lot No:	SM664-B124810
Storage Conditions:	-20°C
Component Lot No:	ROX-224101A
Expiry date:	February 2026

Quality Control Parameters

Analysis	Specification	Result
ROX concentration	A fluorescence spectrophotometer is used to quantify the ROX concentration \pm 5% variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase I.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection 9.7×10^{-3} ng/ μ L RNase.	Passed

QA / QC Representative:



Jan Rahnenführer

 Date: 30th January 2024

United Kingdom

 Tel: +44 (0)20 8830 5300
 Fax: +44 (0)20 8452 2822

USA

 Tel: +1 901.382.8716
 Fax: +1 901.382.0027

Germany

 Tel: +49 (0)3371 60222 00
 Fax: +49 (0)3371 60222 01