

dUTP Mix 50mM

Suitable for Research and further Manufacturing Use

Catalog No:	BIO-39041
Lot No:	DM414-B106870
Storage Conditions:	-20°C
Component Lot No:	DUM-122104A
Expiry date:	May 2024

Quality Control Parameters

Ultra-pure (>99% by HPLC) dNTPs supplied as lithium salts

Analysis	Specification	Result
Functional	A 800bp human genomic DNA fragment is amplified with a dilution series of dUTPs, using standard conditions and 30 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection 9.7×10^{-3} ng/ μ L RNase.	Passed
Nicking Activity	Incubation of dUTP mix with supercoiled control plasmid. Analysed by agarose gel electrophoresis. Test sample does not show an increase of linearized or relaxed plasmid.	Passed

QA / QC Representative:



Alberta Newton

Date: 14th April 2022

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01

Australia

Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763