	Certificate of Analysis	COA No: CA BMM-0006
		Version: 09

MangoMix™ For research or further manufacturing use only	Catalog No:	BIO-25033
	Lot No:	PM335-B122650
	Storage Conditions:	-20°C
	Component Lot No:	BLMX-323211A
	Expiry date:	December 2025

Quality Control Parameters

Analysis	Specification	Result
Functional	A 0.9Kb fragment is amplified with a dilution series of human genomic DNA, using standard conditions and 30 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase.	Passed

QA / QC Representative:



Andrew Galeeba-M

Date: 18th December 2023

United Kingdom


Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01

	Certificate of Analysis	COA No: CA_XBB-0014
		Version: 09

MgCl₂ Solution, 50mM For research or further manufacturing use only	Catalog No:	BIO-25033
	Lot No:	PM335-B122650
	Storage Conditions:	-20°C
	Component Lot No:	MG-2031.017
	Expiry date:	December 2025

Quality Control Parameters

Analysis	Specification	Result
Functional	Fragments of sizes 800bp and 3000bp are amplified with a dilution series of BIOTAQ™ DNA Polymerase, using standard conditions and 30 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase.	Passed

QA / QC Representative:



Andrew Galeeba-M

Date: 18th December 2023

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01