

## Certificate of Analysis

COA No: CA\_BSM-0007

Version: 09

# SensiFAST™ SYBR® & Fluorescein Kit

For research or further manufacturing use only

Catalog No: BIO-96020

Lot No: SF585-B126120

Storage Conditions: -20°C


Component Lot No: SFSF-324103A

Expiry date: April 2026

## Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative PCR analysis amplifying 6 genes from a dilution series of mouse cDNA under standard conditions. Cq and melting profiles must be consistent for the test and reference sample with $\pm 0.5$ Cq variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase I.	Passed

QA / QC Representative:



J. Rahnenführer

Date: 14<sup>th</sup> March 2024

### United Kingdom

Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

### USA

Tel: +1 901.382.8716  
Fax: +1 901.382.0027

### Germany

Tel: +49 (0)3371 60222 00  
Fax: +49 (0)3371 60222 01