

SensiPLUS™ No-ROX Mix 2X

For research or further manufacturing use only

Catalog No:	BIO-11021
Lot No:	SF845-B114980
Storage Conditions:	-20°C
Component Lot No:	SPN-323103A
Expiry date:	April 2025

Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative PCR analysis amplifying 6 genes from a dilution series of mouse cDNA under standard conditions. Cq profiles must be consistent for the test and reference sample with ± 0.5 Cq variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in concordance with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must exhibit less degradation than the limit of detection 2.5 x 10 ⁻³ U DNase.	Passed

QA / QC Representative:



Andrew Galeeba-M

Date: 8th March 2023

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01

Australia

Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763