

HyperLadder™ 25bp

For research or further manufacturing use only

Catalog No:	BIO-33032
Lot No:	MW435-B131310
Storage Conditions:	-20°C
Component Lot No:	H5-324109A
Expiry date:	October 2026

Quality Control Parameters

Certified Values:

Number of Bases	Method of Testing	Specification	Method of Testing	Results
25 bp	Sequencing	80 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
50 bp	Sequencing	80ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
75 bp	Sequencing	80 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
100 bp	Sequencing	120 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
125 bp	Sequencing	80 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
150 bp	Sequencing	80 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
175 bp	Sequencing	80 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
200 bp	Sequencing	120 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
250 bp	Sequencing	60 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
300 bp	Sequencing	60 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027


Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01

400 bp	Sequencing	60 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
500 bp	Sequencing	60 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed

Note: The values given relate to individual bands. Following the combination of all bands in one solution, the Ladder may be used for approximating the mass of DNA.

QA / QC Representative:



J. Rahnenführer

 Date: 27th September 2024

United Kingdom

 Tel: +44 (0)20 8830 5300
 Fax: +44 (0)20 8452 2822

USA

 Tel: +1 901.382.8716
 Fax: +1 901.382.0027

Germany

 Tel: +49 (0)3371 60222 00
 Fax: +49 (0)3371 60222 01

DNA Loading Buffer Blue

For research or further manufacturing use only

Catalog No:	BIO-33032
Lot No:	MW435-B131310
Storage Conditions:	-20°C
Component Lot No:	HLBB-2035.018
Expiry date:	October 2026

Quality Control Parameters

Analysis	Specification	Result
Functional	Tested on a 1.5% gel with 4 different sized DNA. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase.	Passed

QA / QC Representative:



J. Rahnenführer

Date: 27th September 2024

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01