

SensiFAST™ Probe Lo-ROX One-Step Kit

Suitable for Research and further Manufacturing Use as an IVD component

Catalog No:	BIO-78001
Lot No:	SF622-B081200
Shipping / Storage Conditions:	-20°C
Component Lot No:	SFPL1S-920203A
Expiry date:	April 2022

Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative PCR analysis amplifying 6 genes from a dilution series of mouse RNA under standard conditions. Cq profiles must be consistent for the test and reference sample with ± 0.5 Cq variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase I.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection 9.7×10^{-3} ng/μL RNase.	Passed

Authorised by Christopher Weatherall



RNase Inhibitor

Suitable for Research and further Manufacturing Use as an IVD component

Catalog No:	BIO-78001
Lot No:	SF622-B081200
Shipping / Storage Conditions:	-20°C
Component Lot No:	RI-920203A
Expiry date:	April 2022

Quality Control Parameters

Analysis	Specification	Result
Inhibition	Test level of inhibition by incubating total RNA with concentration gradient of RNase A. Bands were observed with agarose gel electrophoresis (ethidium stained).	Passed

Authorised by Christopher Weatherall



Reverse Transcriptase

Suitable for Research and further Manufacturing Use as an IVD component

Catalog No:	BIO-78001
Lot No:	SF622-B081200
Shipping / Storage Conditions:	-20°C
Component Lot No:	RTP-920303A
Expiry date:	April 2022

Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative PCR analysis amplifying 6 genes from a dilution series of mouse RNA under standard conditions. Cq profiles must be consistent for the test and reference sample with ± 0.5 Cq variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase I.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection 9.7×10^{-3} ng/ μ L RNase.	Passed

Authorised by Christopher Weatherall



DEPC Water

Suitable for Research and further Manufacturing Use as an IVD component

Catalog No:	BIO-78001
Lot No:	SF622-B081200
Shipping / Storage Conditions:	-20°C
Component Lot No:	DWT-819109A
Expiry date:	April 2022

Quality Control Parameters

Analysis	Specification	Result
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase I.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection 9.7×10^{-3} ng/ μ L RNase.	Passed

Authorised by Christopher Weatherall

