

## MyFi™ DNA Polymerase

For research or further manufacturing use only

Catalog No:	BIO-21119
Lot No:	PL368-B110240
Storage Conditions:	-20°C
Component Lot No:	MF-122208A
Expiry date:	September 2024

### Quality Control Parameters

Analysis	Specification	Result
Functional	Fragment of size 525bp is amplified with a dilution series of MyFi DNA Polymerase, using standard conditions and 35 cycles. Fragments of sizes 7Kb and 1300bp are amplified with a dilution series of human genomic DNA, using standard conditions and 35 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase.	Passed

QA / QC Representative:



Alberta Newton

 Date: 10<sup>th</sup> August 2022

**United Kingdom**

 Tel: +44 (0)20 8830 5300  
 Fax: +44 (0)20 8452 2822

**USA**

 Tel: +1 901.382.8716  
 Fax: +1 901.382.0027

**Germany**

 Tel: +49 (0)3371 60222 00  
 Fax: +49 (0)3371 60222 01

**Australia**

 Tel: +61 (0)2 9209 4180  
 Fax: +61 (0)2 9209 4763

**MyFi™ Buffer**

For research or further manufacturing use only

Catalog No:	BIO-21119
Lot No:	PL368-B110240
Storage Conditions:	-20°C
Component Lot No:	MFB-222208A
Expiry date:	September 2024

**Quality Control Parameters**

Analysis	Specification	Result
Functional	Fragments of sizes 525bp and 7Kb were amplified with a dilution series of human genomic DNA, using standard conditions and 35 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase.	Passed

QA / QC Representative:



Alberta Newton

Date: 10<sup>th</sup> August 2022

**United Kingdom**

Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**USA**

Tel: +1 901.382.8716  
Fax: +1 901.382.0027

**Germany**

Tel: +49 (0)3371 60222 00  
Fax: +49 (0)3371 60222 01

**Australia**

Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763