

ISOLATE II Gel and PCR Kit

Suitable for Research and further Manufacturing Use

Catalog No:	BIO-52060
Lot No:	IS497-B103060
Storage Conditions:	Room Temperature
Component Lot No:	IS497-B103060
Expiry date:	12 months from date of purchase

Quality Control Parameters

Component	Specification	Result	Batch
ISOLATE II PCR and Gel Columns	Nucleic acid recovery >90%	Passed	SPI002784
Collection tubes	n/a	n/a	COL001889
Binding Buffer CB	pH = 4.8-5.2, density	Passed	103572
Wash Buffer CW (conc)	pH = 7.5-7.9, density	Passed	103571
Elution Buffer C	pH = 8.3-8.7, density	Passed	100096

QA / QC Representative:



Andrew Galeeba-M

Date: 14th February 2022

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01

Australia

Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763