

ISOLATE II RNA Mini Kit

Suitable for Research and further Manufacturing Use

Catalog No:	BIO-52073
Lot No:	IS510-B098230
Storage Conditions:	Room Temperature
Component Lot No:	IS510-B098230
Expiry date:	12 months from date of purchase

Quality Control Parameters

Component	Specification	Result	Batch
ISOLATE II Filters	n/a	n/a	SPI002819
ISOLATE II RNA Mini Columns	Nucleic acid recovery >90%	Passed	SPB002816
Collection tubes (2 mL)	n/a	n/a	COL001889
Collection tubes (1.5 mL)	n/a	n/a	100221
Lysis Buffer RLY	pH = 6.7-7.3, density	Passed	100521
Wash Buffer RW1	pH = 6.9-7.5, density	Passed	100524
Wash Buffer RW2	pH = 7.5-7.9, density, sterility	Passed	98351
Membrane Desalting Buffer MEM	pH = 7.8-7.1, density	Passed	100519
Reaction Buffer for DNase I RDN	pH = 6.9-7.1, density	Passed	100501
DNase, RNase-free	No RNase activity detectable in connection with our kit	Passed	D2115R
RNase-free Water	Density, no RNase activity, sterility	Passed	100527

QA / QC Representative:



Andrew Galeeba-M

 Date: 19th October 2021

United Kingdom

 Tel: +44 (0)20 8830 5300
 Fax: +44 (0)20 8452 2822

USA

 Tel: +1 901.382.8716
 Fax: +1 901.382.0027

Germany

 Tel: +49 (0)3371 60222 00
 Fax: +49 (0)3371 60222 01

Australia

 Tel: +61 (0)2 9209 4180
 Fax: +61 (0)2 9209 4763