
	<b>Certificate of Analysis</b>	COA No: CA_BEM-0025
		Version: 09

<b>SensiFAST™ Reverse Transcriptase</b>  For research or further manufacturing use only	Catalog No:	BIO-65054
	Lot No:	RA653-B124710
	Storage Conditions:	-20°C
	Component Lot No:	SRT-424102A
	Expiry date:	March 2026

### Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative PCR analysis amplifying 4 genes from a dilution series of mouse RNA under standard conditions. Cq and melting profiles must be consistent for the test and reference sample with $\pm 0.5$ Cq variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase I.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection $9.7 \times 10^{-3}$ ng/ $\mu$ L RNase.	Passed

QA / QC Representative:



J. Rahnenführer

Date: 01<sup>st</sup> March 2024

#### United Kingdom


Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

#### USA

Tel: +1 901.382.8716  
Fax: +1 901.382.0027

#### Germany

Tel: +49 (0)3371 60222 00  
Fax: +49 (0)3371 60222 01


	<b>Certificate of Analysis</b>	COA No: CA_XBB-0050
		Version: 08

<b>5x TransAmp Buffer</b>  For research or further manufacturing use only	Catalog No:	BIO-65054
	Lot No:	RA653-B124710
	Storage Conditions:	-20°C
	Component Lot No:	TAB-424102A
	Expiry date:	March 2026

### Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative PCR analysis amplifying 4 genes from a dilution series of mouse cDNA under standard conditions. cDNA was synthesised using the SensiFAST cDNA synthesis kit, using recommended conditions. Cq and melting profiles for the test must be within the reference variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase I.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection $9.7 \times 10^{-3}$ ng/ $\mu$ L RNase.	Passed

QA / QC Representative:



J. Rahnenführer

Date: 01<sup>st</sup> March 2024

#### United Kingdom

Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

#### USA

Tel: +1 901.382.8716  
Fax: +1 901.382.0027

#### Germany

Tel: +49 (0)3371 60222 00  
Fax: +49 (0)3371 60222 01