

SensiFAST™ Reverse Transcriptase


For research or further manufacturing use only

Catalog No:	BIO-65053
Lot No:	RA652-B128890
Storage Conditions:	-20°C
Component Lot No:	SRT-424207A
Expiry date:	August 2026

Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative PCR analysis amplifying 4 genes from a dilution series of mouse RNA under standard conditions. Cq and melting profiles must be consistent for the test and reference sample with ± 0.5 Cq variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase I.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection 9.7×10^{-3} ng/ μ L RNase.	Passed

QA / QC Representative:



J. Rahnenführer

Date: 11th July 2024

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01

5x TransAmp Buffer


For research or further manufacturing use only

Catalog No:	BIO-65053
Lot No:	RA652-B128890
Storage Conditions:	-20°C
Component Lot No:	TAB-424107A
Expiry date:	August 2026

Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative PCR analysis amplifying 4 genes from a dilution series of mouse cDNA under standard conditions. cDNA was synthesised using the SensiFAST cDNA synthesis kit, using recommended conditions. Cq and melting profiles for the test must be within the reference variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase I.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection 9.7×10^{-3} ng/ μ L RNase.	Passed

QA / QC Representative:



J. Rahnenführer

Date: 11th July 2024

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01