

**SensiFAST™ Probe Hi-ROX One-Step Kit**


For research or further manufacturing use only

Catalog No:	BIO-77001
Lot No:	SF620-B124350
Storage Conditions:	-20°C
Component Lot No:	SFPH1S-224101A
Expiry date:	February 2026

**Quality Control Parameters**

Analysis	Specification	Result
Functional	Quantitative PCR analysis amplifying 6 genes from a dilution series of mouse RNA under standard conditions. Cq profiles must be consistent for the test and reference sample with ± 0.5 Cq variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5 x 10 <sup>-3</sup> U DNase I.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection 9.7x10 <sup>-3</sup> ng/μL RNase.	Passed

QA / QC Representative:



J. Rahnenführer

Date: 17<sup>th</sup> January 2024

**United Kingdom**

Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**USA**

Tel: +1 901.382.8716  
Fax: +1 901.382.0027

**Germany**

Tel: +49 (0)3371 60222 00  
Fax: +49 (0)3371 60222 01

## Reverse Transcriptase


For research or further manufacturing use only

Catalog No:	BIO-77001
Lot No:	SF620-B124350
Storage Conditions:	-20°C
Component Lot No:	RTP-124201A
Expiry date:	February 2026

### Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative PCR analysis amplifying 6 genes from a dilution series of mouse RNA under standard conditions. Cq profiles must be consistent for the test and reference sample with $\pm 0.5$ Cq variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase I.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection $9.7 \times 10^{-3}$ ng/ $\mu$ L RNase.	Passed

QA / QC Representative:



J. Rahnenführer

 Date: 17<sup>th</sup> January 2024

United Kingdom

 Tel: +44 (0)20 8830 5300  
 Fax: +44 (0)20 8452 2822

USA

 Tel: +1 901.382.8716  
 Fax: +1 901.382.0027

Germany

 Tel: +49 (0)3371 60222 00  
 Fax: +49 (0)3371 60222 01

**RNase Inhibitor**


Suitable for Research and further Manufacturing Use

Catalog No:	BIO-77001
Lot No:	SF620-B124350
Storage Conditions:	-20°C
Component Lot No:	RI-124301A
Expiry date:	February 2026

**Quality Control Parameters**

Analysis	Specification	Result
Inhibition	Test level of inhibition by incubating total RNA with concentration gradient of RNase A. Bands were observed with agarose gel electrophoresis (ethidium stained).	Passed

QA / QC Representative:



J. Rahnenführer

Date: 17<sup>th</sup> January 2024

**United Kingdom**

Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**USA**

Tel: +1 901.382.8716  
Fax: +1 901.382.0027

**Germany**

Tel: +49 (0)3371 60222 00  
Fax: +49 (0)3371 60222 01

**DEPC Water**

For research or further manufacturing use only

Catalog No:	BIO-77001
Lot No:	SF620-B124350
Storage Conditions:	-20°C
Component Lot No:	DWT-124901B
Expiry date:	February 2026

**Quality Control Parameters**

Analysis	Specification	Result
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase I.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection $9.7 \times 10^{-3}$ ng/ $\mu$ L RNase.	Passed

QA / QC Representative:



J. Rahnenführer

Date: 17<sup>th</sup> January 2024

**United Kingdom**

Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**USA**

Tel: +1 901.382.8716  
Fax: +1 901.382.0027

**Germany**

Tel: +49 (0)3371 60222 00  
Fax: +49 (0)3371 60222 01