

MyTaq™ Blood PCR Mix


For research or further manufacturing use only

| | |
|---------------------|---------------|
| Catalog No: | BIO-25054 |
| Lot No: | PM642-B126570 |
| Storage Conditions: | -20°C |
| Component Lot No: | MTBX-224104A |
| Expiry date: | May 2026 |

Quality Control Parameters

| Analysis | Specification | Result |
|---------------------|---|--------|
| Functional | <p>Fragment of size 5Kb was amplified with a dilution series human blood, using standard conditions and 40 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).</p> <p>A four plex PCR reaction was run using standard conditions, 40 cycles with a dilution series of mix and 35 cycles with a dilution series of human blood. Bands sizes were 1.2Kb, 844bp, 450bp and 237bp. Distinct bands were observed with agarose gel electrophoresis (ethidium stained).</p> | Passed |
| DNA contamination | Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample. | Passed |
| DNase contamination | Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase. | Passed |

QA / QC Representative:



J. Rahnenführer

 Date: 11th April 2024

United Kingdom

 Tel: +44 (0)20 8830 5300
 Fax: +44 (0)20 8452 2822

USA

 Tel: +1 901.382.8716
 Fax: +1 901.382.0027

Germany

 Tel: +49 (0)3371 60222 00
 Fax: +49 (0)3371 60222 01