

## MyTaq™ Blood PCR Mix

For research or further manufacturing use only

Catalog No:	BIO-25054
Lot No:	PM642-B106960
Storage Conditions:	-20°C
Component Lot No:	MTBX-122104A
Expiry date:	May 2024

### Quality Control Parameters

Analysis	Specification	Result
Functional	<p>Fragment of size 5Kb was amplified with a dilution series human blood, using standard conditions and 40 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).</p> <p>A four plex PCR reaction was run using standard conditions, 40 cycles with a dilution series of mix and 35 cycles with a dilution series of human blood. Bands sizes were 1.2Kb, 844bp, 450bp and 237bp. Distinct bands were observed with agarose gel electrophoresis (ethidium stained).</p>	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase.	Passed

QA / QC Representative:



Andrew Galeeba-M

 Date: 22<sup>nd</sup> April 2022

**United Kingdom**

 Tel: +44 (0)20 8830 5300  
 Fax: +44 (0)20 8452 2822

**USA**

 Tel: +1 901.382.8716  
 Fax: +1 901.382.0027

**Germany**

 Tel: +49 (0)3371 60222 00  
 Fax: +49 (0)3371 60222 01

**Australia**

 Tel: +61 (0)2 9209 4180  
 Fax: +61 (0)2 9209 4763