

EasyLadder I

For research or further manufacturing use only

Catalog No:	BIO-33046
Lot No:	MW438-B101370
Storage Conditions:	-20°C
Component Lot No:	HTS1-021111A
Expiry date:	December 2023

Quality Control Parameters

Certified Values:

Number of Bases	Method of Testing	Specification	Method of Testing	Results
100 bp	Sequencing	50 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
250 bp	Sequencing	50 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
500 bp	Sequencing	50 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
1000 bp	Sequencing	50 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed
2000 bp	Sequencing	50 ng/band ± 10%	UV absorption spectrum Visual comparison test vs history	Passed

Note: The values given relate to individual bands. Following the combination of all bands in one solution, the Ladder may be used for approximating the mass of DNA.

QA / QC Representative:



Alberta Newton

 Date: 16th November 2021

United Kingdom

 Tel: +44 (0)20 8830 5300
 Fax: +44 (0)20 8452 2822

USA

 Tel: +1 901.382.8716
 Fax: +1 901.382.0027

Germany

 Tel: +49 (0)3371 60222 00
 Fax: +49 (0)3371 60222 01

Australia

 Tel: +61 (0)2 9209 4180
 Fax: +61 (0)2 9209 4763

DNA Loading Buffer Red

For research or further manufacturing use only

Catalog No:	BIO-33046
Lot No:	MW438-B101370
Storage Conditions:	-20°C
Component Lot No:	HLBR-921111A
Expiry date:	December 2023

Quality Control Parameters

Analysis	Specification	Result
Functional	Tested on a 1.5% gel with 4 different sized DNA. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase.	Passed

QA / QC Representative:



Alberta Newton

Date: 16th November 2021

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01

Australia

Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763