



Certificate of Analysis

COA No: CA_RES-0020

Version: 01

ISOLATE II Gel and PCR Kit

For Research Use Only

Storage Conditions:	Room Temperature
Lot number:	IS495:313H
Expiry date:	January 2015

Quality Control Parameters

Component	Specification	Result	Batch
ISOLATE II PCR and Gel Columns	Nucleic acid recovery >90%	Passed	SC-313H
Collection tubes	n/a	n/a	OT-313H
Binding Buffer CB	pH = 4.8-5.2, density	Passed	CB-313H
Wash Buffer CW (conc)	pH = 7.5-7.9, density	Passed	CW-313H
Elution Buffer C	pH = 8.3-8.7, density	Passed	EBC-313H

Authorised by Jade James

Europe
Headquarters UK

info.uk@bioline.com
Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

Europe
Germany

info.de@bioline.com
Tel: +49 (0)3371 681 229
Fax: +49 (0)3371 681 244

America

info.us@bioline.com
Tel: +1 508 880 8990
Fax: +1 508 880 8993

Australia

info.aust@bioline.com
Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763