



## Certificate of Analysis

COA No: CA BB-0003

Version: 01

### RT Buffer

For Research Use Only

Storage Conditions: -20°C

Lot number: TRTB-313209

Expiry date: September 2015

### Quality Control Parameters

Analysis	Specification	Result
Functional	Fragment of size 1Kb was reverse transcribed with Bioscript, with a template dilution series, using standard conditions. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection $9.7 \times 10^{-3}$ ng/ $\mu$ l RNase.	Passed

Authorised by Jade James

**Europe**  
Headquarters UK

[info.uk@bioline.com](mailto:info.uk@bioline.com)  
Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**Europe**  
Germany

[info.de@bioline.com](mailto:info.de@bioline.com)  
Tel: +49 (0)3371 681 229  
Fax: +49 (0)3371 681 244

**America**

[info.us@bioline.com](mailto:info.us@bioline.com)  
Tel: +1 508 880 8990  
Fax: +1 508 880 8993

**Australia**

[info.aust@bioline.com](mailto:info.aust@bioline.com)  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763



## Certificate of Analysis

COA No: CA CHM-0192

Version: 01

### Reverse Transcriptase

For Research Use Only

Storage Conditions: -20°C

Lot number: TRT-213209

Expiry date: September 2015

### Quality Control Parameters

Analysis	Specification	Result
Functional	Fragments of sizes 1.2Kb and 6.5Kb were reverse transcribed, using standard conditions. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
Endonuclease contamination	Super coiled DNA plasmid was incubated with the reverse transcriptase for 1 hour at 37°C, the absence of nicking and cutting is shown by agarose gel electrophoresis.	Passed
DNase and RNase contamination	A DNA and RNA fragment were Incubated with the reverse transcriptase for 1 hour at 37°C. < 1% degradation was observed.	Passed

Authorised by Jade James

**Europe**  
Headquarters UK

[info.uk@bioline.com](mailto:info.uk@bioline.com)  
Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**Europe**  
Germany

[info.de@bioline.com](mailto:info.de@bioline.com)  
Tel: +49 (0)3371 681 229  
Fax: +49 (0)3371 681 244

**America**

[info.us@bioline.com](mailto:info.us@bioline.com)  
Tel: +1 508 880 8990  
Fax: +1 508 880 8993

**Australia**

[info.aust@bioline.com](mailto:info.aust@bioline.com)  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763



## Certificate of Analysis

COA No: CA BE-0031

Version: 01

### RNase Inhibitor

For Research Use Only

Storage Conditions: -20°C

Lot number: RI-113109

Expiry date: September 2015

### Quality Control Parameters

Analysis	Specification	Result
Inhibition	Test level of inhibition by incubating total RNA with concentration gradient of RNase A. Bands were observed with agarose gel electrophoresis (ethidium stained).	Passed

Authorised by Jade James

**Europe**  
Headquarters UK

[info.uk@bioline.com](mailto:info.uk@bioline.com)  
Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**Europe**  
Germany

[info.de@bioline.com](mailto:info.de@bioline.com)  
Tel: +49 (0)3371 681 229  
Fax: +49 (0)3371 681 244

**America**

[info.us@bioline.com](mailto:info.us@bioline.com)  
Tel: +1 508 880 8990  
Fax: +1 508 880 8993

**Australia**

[info.aust@bioline.com](mailto:info.aust@bioline.com)  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763



## Certificate of Analysis

COA No: CA BNN-0003

Version: 01

### dNTP Mix 10mM

For Research Use Only

Storage Conditions: -20°C

Lot number: DM1-313109

Expiry date: September 2015

### Quality Control Parameters

Analysis	Specification	Result
Functional	A 800bp human genomic DNA fragment is amplified with a dilution series of enzymes, using standard conditions and 30 cycles. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with a reference sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection $9.7 \times 10^{-3}$ ng/ $\mu$ l RNase.	Passed

Authorised by Jade James

**Europe**  
Headquarters UK

[info.uk@bioline.com](mailto:info.uk@bioline.com)  
Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**Europe**  
Germany

[info.de@bioline.com](mailto:info.de@bioline.com)  
Tel: +49 (0)3371 681 229  
Fax: +49 (0)3371 681 244

**America**

[info.us@bioline.com](mailto:info.us@bioline.com)  
Tel: +1 508 880 8990  
Fax: +1 508 880 8993

**Australia**

[info.aust@bioline.com](mailto:info.aust@bioline.com)  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763



## Certificate of Analysis

COA No: CA BS-0085

Version: 01

### Oligo (dT)<sub>18</sub>

For Research Use Only

Storage Conditions: -20°C

Lot number: ODT-013109

Expiry date: September 2015

### Quality Control Parameters

Analysis	Specification	Result
Functional	A 1Kb fragment is reverse transcribed from Poly A RNA with a dilution series of Bioscript, using standard conditions. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection $9.7 \times 10^{-3}$ ng/ $\mu$ l RNase.	Passed

Authorised by Jade James

**Europe**  
Headquarters UK

[info.uk@bioline.com](mailto:info.uk@bioline.com)  
Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**Europe**  
Germany

[info.de@bioline.com](mailto:info.de@bioline.com)  
Tel: +49 (0)3371 681 229  
Fax: +49 (0)3371 681 244

**America**

[info.us@bioline.com](mailto:info.us@bioline.com)  
Tel: +1 508 880 8990  
Fax: +1 508 880 8993

**Australia**

[info.aust@bioline.com](mailto:info.aust@bioline.com)  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763



## Certificate of Analysis

COA No: CA BS-0090

Version: 01

### Random Hexamer Primer

For Research Use Only

Storage Conditions: -20°C

Lot number: RHP-113209

Expiry date: September 2015

### Quality Control Parameters

Analysis	Specification	Result
Functional	A 1Kb fragment is reverse transcribed from Poly A RNA with a dilution series of Bioscript, using standard conditions. Single distinct bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection $9.7 \times 10^{-3}$ ng/ $\mu$ l RNase.	Passed

Authorised by Jade James

**Europe**  
Headquarters UK

[info.uk@bioline.com](mailto:info.uk@bioline.com)  
Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**Europe**  
Germany

[info.de@bioline.com](mailto:info.de@bioline.com)  
Tel: +49 (0)3371 681 229  
Fax: +49 (0)3371 681 244

**America**

[info.us@bioline.com](mailto:info.us@bioline.com)  
Tel: +1 508 880 8990  
Fax: +1 508 880 8993

**Australia**

[info.aust@bioline.com](mailto:info.aust@bioline.com)  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763



## Certificate of Analysis

COA No: CA\_BS-0020

Version: 01

### DEPC Water

For Research Use Only

Storage  
Conditions:

-20°C

Lot number:

DWT-213109

Expiry date:

September 2015

### Quality Control Parameters

Analysis	Specification	Result
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in line with control sample.	Passed
DNase contamination	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection $2.5 \times 10^{-3}$ U DNase I.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection $9.7 \times 10^{-3}$ ng/ $\mu$ l RNase.	Passed

Authorised by Jade James

**Europe**  
Headquarters UK

[info.uk@bioline.com](mailto:info.uk@bioline.com)  
Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**Europe**  
Germany

[info.de@bioline.com](mailto:info.de@bioline.com)  
Tel: +49 (0)3371 681 229  
Fax: +49 (0)3371 681 244

**America**

[info.us@bioline.com](mailto:info.us@bioline.com)  
Tel: +1 508 880 8990  
Fax: +1 508 880 8993

**Australia**

[info.aust@bioline.com](mailto:info.aust@bioline.com)  
Tel: +61 (0)2 9209 4180  
Fax: +61 (0)2 9209 4763